

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR080218CLP.M
 Title : GC EXTRACTABLES
 Last Update : Thu Aug 02 02:04:29 2018
 Response Via : Initial Calibration

Calibration Files

400 =PR031049.D 1600 =PR031051.D 800 =PR031050.D
 200 =PR031048.D 100 =PR031047.D

	Compound	400	1600	800	200	100	Avg	%RSD
1) SA	Tetrachloro-m-xylene	1.701	1.273	1.537	1.518	1.805	1.567 E7	12.94
2) SA	Decachlorobiphenyl	3.575	3.181	3.247	3.886	4.330	3.644 E7	13.06
3) L1	AR-1016-1						0.000	-1.00
4) L1	AR-1016-2						0.000	-1.00
5) L1	AR-1016-3						0.000	-1.00
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1	1.543	1.321	1.364	1.562	1.811	1.520 E6	12.77
27) L6	AR-1254-2	0.951	0.831	0.847	0.948	1.070	0.929 E6	10.36
28) L6	AR-1254-3	1.729	1.525	1.549	1.747	1.998	1.710 E6	11.14
29) L6	AR-1254-4	1.518	1.334	1.347	1.550	1.787	1.507 E6	12.22
30) L6	AR-1254-5	1.091	0.981	0.975	1.106	1.223	1.075 E6	9.50
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

Signal #2 Calibration Files

400 =PR031049.D 1600 =PR031051.D 800 =PR031050.D
 200 =PR031048.D 100 =PR031047.D

	Compound	400	1600	800	200	100	Avg	%RSD
1) SA	Tetrachloro-m-xylene	2.865	2.130	2.569	2.582	3.105	2.650 E7	13.79
2) SA	Decachlorobiphenyl	2.202	1.867	1.956	2.466	2.780	2.254 E7	16.63
3) L1	AR-1016-1						0.000	-1.00
4) L1	AR-1016-2						0.000	-1.00
5) L1	AR-1016-3						0.000	-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR080218CLP.M
 Title : GC EXTRACTABLES
 Last Update : Thu Aug 02 02:04:29 2018
 Response Via : Initial Calibration

Calibration Files

400 =PR031049.D 1600 =PR031051.D 800 =PR031050.D
 200 =PR031048.D 100 =PR031047.D

	Compound	400	1600	800	200	100	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1	2.735	2.267	2.386	2.769	3.294	2.690 E6	14.93
27) L6	AR-1254-2	2.281	1.937	1.964	2.301	2.673	2.231 E6	13.44
28) L6	AR-1254-3	3.586	3.021	3.130	3.691	4.332	3.552 E6	14.69
29) L6	AR-1254-4	1.860	1.577	1.628	1.930	2.264	1.852 E6	14.84
30) L6	AR-1254-5	3.491	2.944	2.974	3.644	4.235	3.458 E6	15.43
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range