

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR082018.M
 Title : GC EXTRACTABLES
 Last Update : Tue Aug 21 07:29:33 2018
 Response Via : Initial Calibration

Calibration Files

500 =PR031492.D 1000 =PR031490.D 750 =PR031491.D
 250 =PR031493.D 50 =PR031494.D

Compound			500	1000	750	250	50	Avg	%RSD	
1)	SA	Tetrachloro-m-xylene	0.980	1.240	0.729	1.142	0.983	1.015	E7	19.15
2)	SA	Decachlorobiphenyl	3.004	3.088	2.785	3.259	3.134	3.054	E7	5.77
3)	L1	AR-1016-1	1.963	2.364	1.603	2.364	1.912	2.041	E5	15.94
4)	L1	AR-1016-2	3.130	3.879	2.716	3.639	2.780	3.229	E5	15.98
5)	L1	AR-1016-3	7.232	8.801	6.353	8.366	7.425	7.635	E5	12.67
6)	L1	AR-1016-4	4.581	5.500	4.082	5.181	4.509	4.770	E5	11.86
7)	L1	AR-1016-5	4.473	5.443	4.090	4.967	4.421	4.679	E5	11.33
8)	L2	AR-1221-1	9.125					9.125	E4	0.00
9)	L2	AR-1221-2	7.230					7.230	E4	0.00
10)	L2	AR-1221-3	2.291					2.291	E5	0.00
11)	L3	AR-1232-1	1.100					1.100	E5	0.00
12)	L3	AR-1232-2	1.197					1.197	E5	0.00
13)	L3	AR-1232-3	2.771					2.771	E5	0.00
14)	L3	AR-1232-4	1.779					1.779	E5	0.00
15)	L3	AR-1232-5	1.428					1.428	E5	0.00
16)	L4	AR-1242-1	1.771					1.771	E5	0.00
17)	L4	AR-1242-2	2.872					2.872	E5	0.00
18)	L4	AR-1242-3	4.050					4.050	E5	0.00
19)	L4	AR-1242-4	2.381					2.381	E5	0.00
20)	L4	AR-1242-5	2.685					2.685	E5	0.00
21)	L5	AR-1248-1	4.165					4.165	E5	0.00
22)	L5	AR-1248-2	3.548					3.548	E5	0.00
23)	L5	AR-1248-3	5.516					5.516	E5	0.00
24)	L5	AR-1248-4	5.270					5.270	E5	0.00
25)	L5	AR-1248-5	5.645					5.645	E5	0.00
26)	L6	AR-1254-1	3.058					3.058	E5	0.00
27)	L6	AR-1254-2	1.097					1.097	E6	0.00
28)	L6	AR-1254-3	1.314					1.314	E6	0.00
29)	L6	AR-1254-4	1.117					1.117	E6	0.00
30)	L6	AR-1254-5	1.168					1.168	E6	0.00
31)	L7	AR-1260-1	1.031	1.187	0.950	1.150	1.071	1.078	E6	8.74
32)	L7	AR-1260-2	1.413	1.631	1.305	1.592	1.442	1.477	E6	9.09
33)	L7	AR-1260-3	1.667	1.902	1.552	1.884	1.610	1.723	E6	9.30
34)	L7	AR-1260-4	3.070	3.462	2.878	3.299	2.998	3.141	E6	7.51
35)	L7	AR-1260-5	1.967	2.149	1.823	2.169	2.054	2.032	E6	6.99
36)	L8	AR-1262-1	9.296					9.296	E5	0.00
37)	L8	AR-1262-2	1.234					1.234	E6	0.00
38)	L8	AR-1262-3	1.897					1.897	E6	0.00
39)	L8	AR-1262-4	1.634					1.634	E6	0.00
40)	L8	AR-1262-5	3.764					3.764	E6	0.00
41)	L9	AR-1268-1	3.949					3.949	E6	0.00
42)	L9	AR-1268-2	3.493					3.493	E6	0.00
43)	L9	AR-1268-3	3.163					3.163	E6	0.00
44)	L9	AR-1268-4	1.461					1.461	E6	0.00
45)	L9	AR-1268-5	9.441					9.441	E6	0.00

Signal #2 Calibration Files

500 =PR031492.D 1000 =PR031490.D 750 =PR031491.D
 250 =PR031493.D 50 =PR031494.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	1.659	2.089	1.261	1.940	1.728	1.735	E7	18.17
2) SA	Decachlorobiphenyl	2.040	2.050	1.850	2.270	2.250	2.092	E7	8.26
3) L1	AR-1016-1	5.642	6.560	4.697	6.632	6.414	5.989	E5	13.74
4) L1	AR-1016-2	0.860	1.031	0.755	0.981	0.863	0.898	E6	12.19
5) L1	AR-1016-3	1.550	1.891	1.379	1.862	1.619	1.660	E6	13.02

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR082018.M
 Title : GC EXTRACTABLES
 Last Update : Tue Aug 21 07:29:33 2018
 Response Via : Initial Calibration

Calibration Files

500 =PR031492.D 1000 =PR031490.D 750 =PR031491.D
 250 =PR031493.D 50 =PR031494.D

Compound			500	1000	750	250	50	Avg		%RSD
6)	L1	AR-1016-4	1.002	1.166	0.914	1.129	0.860	1.014	E6	13.07
7)	L1	AR-1016-5	6.490	7.756	5.882	7.464	5.872	6.693	E5	13.15
8)	L2	AR-1221-1	1.592					1.592	E5	0.00
9)	L2	AR-1221-2	1.299					1.299	E5	0.00
10)	L2	AR-1221-3	4.274					4.274	E5	0.00
11)	L3	AR-1232-1	1.370					1.370	E5	0.00
12)	L3	AR-1232-2	6.154					6.154	E5	0.00
13)	L3	AR-1232-3	3.809					3.809	E5	0.00
14)	L3	AR-1232-4	2.760					2.760	E5	0.00
15)	L3	AR-1232-5	3.760					3.760	E5	0.00
16)	L4	AR-1242-1	3.543					3.543	E5	0.00
17)	L4	AR-1242-2	9.051					9.051	E5	0.00
18)	L4	AR-1242-3	4.371					4.371	E5	0.00
19)	L4	AR-1242-4	3.672					3.672	E5	0.00
20)	L4	AR-1242-5	5.272					5.272	E5	0.00
21)	L5	AR-1248-1	6.614					6.614	E5	0.00
22)	L5	AR-1248-2	6.913					6.913	E5	0.00
23)	L5	AR-1248-3	8.942					8.942	E5	0.00
24)	L5	AR-1248-4	1.476					1.476	E6	0.00
25)	L5	AR-1248-5	1.056					1.056	E6	0.00
26)	L6	AR-1254-1	6.326					6.326	E5	0.00
27)	L6	AR-1254-2	1.821					1.821	E6	0.00
28)	L6	AR-1254-3	3.104					3.104	E6	0.00
29)	L6	AR-1254-4	2.466					2.466	E6	0.00
30)	L6	AR-1254-5	2.394					2.394	E6	0.00
31)	L7	AR-1260-1	2.144	2.434	1.991	2.475	2.470	2.303	E6	9.63
32)	L7	AR-1260-2	2.692	3.028	2.504	3.083	3.207	2.903	E6	10.11
33)	L7	AR-1260-3	2.067	2.206	1.896	2.336	2.080	2.117	E6	7.79
34)	L7	AR-1260-4	3.133	3.216	2.827	3.749	3.938	3.373	E6	13.59
35)	L7	AR-1260-5	1.919	1.927	1.714	2.283	2.329	2.035	E6	12.93
36)	L8	AR-1262-1	1.980					1.980	E6	0.00
37)	L8	AR-1262-2	2.393					2.393	E6	0.00
38)	L8	AR-1262-3	2.907					2.907	E6	0.00
39)	L8	AR-1262-4	2.004					2.004	E6	0.00
40)	L8	AR-1262-5	3.633					3.633	E6	0.00
41)	L9	AR-1268-1	2.986					2.986	E6	0.00
42)	L9	AR-1268-2	2.564					2.564	E6	0.00
43)	L9	AR-1268-3	2.058					2.058	E6	0.00
44)	L9	AR-1268-4	8.895					8.895	E5	0.00
45)	L9	AR-1268-5	6.098					6.098	E6	0.00

 (#) = Out of Range