

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR082019\
 Data File : PR040460.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Aug 2019 12:14
 Operator : SM\AJ
 Sample : K3591-07
 Misc : AR1248 LOD 40 PPB
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 LOD-MDL-WATER-01-QT3-2019

Manual Integrations
 APPROVED

Ankita
 8/21/2019 8:10:19 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 21 00:18:40 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR081919.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Aug 20 09:47:16 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.731	4.575	23507275	111.3E6	13.443m	13.288
2)	SA Decachlor...	8.653	10.282	57055408	236.8E6	20.692	18.182
Target Compounds							
21)	L5 AR-1248-1	4.797	5.728	1378772	6167854	32.891m	32.767
22)	L5 AR-1248-2	5.027	5.995	2065442	8256849	35.238m	32.551
23)	L5 AR-1248-3	5.067	6.199	2166545	9714904	35.404m	32.408
24)	L5 AR-1248-4	5.236	6.599	2687449	13868998	34.505	37.495
25)	L5 AR-1248-5	5.621	6.636	2521190	13252697	30.221	37.761

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR082019\
Data File : PR040460.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Aug 2019 12:14
Operator : SM\AJ
Sample : K3591-07
Misc : AR1248 LOD 40 PPB
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
LOD-MDL-WATER-01-QT3-2019

Manual Integrations
APPROVED

Ankita
8/21/2019 8:10:19 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 21 00:18:40 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR081919.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Aug 20 09:47:16 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

