

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR093019\
 Data File : PR041705.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Oct 2019 05:37
 Operator : SM\AJ
 Sample : AR1248CCC400
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR1248356

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 01 06:21:34 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR092319CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 24 04:36:18 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.738	4.004	22079036	95907465	16.633	15.855
2)	SA Decachlor...	10.655	9.115	143.4E6	437.0E6	44.931	40.511
Target Compounds							
21)	L5 AR-1248-1	5.912	5.100	14735410	30855293	401.446	366.070
22)	L5 AR-1248-2	6.184	5.338	21630462	45715536	392.691	375.861
23)	L5 AR-1248-3	6.391	5.381	25560457	48041770	398.075	378.942
24)	L5 AR-1248-4	6.794	5.554	33967650	63607202	398.013	378.960
25)	L5 AR-1248-5	6.833	5.950	32656452	87261492	400.768	400.775

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR093019\
Data File : PR041705.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Oct 2019 05:37
Operator : SM\AJ
Sample : AR1248CCC400
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR1248356

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 01 06:21:34 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR092319CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 24 04:36:18 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

