

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR101018CLP.M
 Title : GC EXTRACTABLES
 Last Update : Thu Oct 11 10:37:15 2018
 Response Via : Initial Calibration

Calibration Files

400 =PR032723.D 1600 =PR032725.D 800 =PR032724.D
 200 =PR032722.D 100 =PR032721.D

Compound			400	1600	800	200	100	Avg	%RSD	
1)	SA	Tetrachloro-m-xylene	1.290	1.423	1.287	1.440	1.554	1.399	E7	8.06
2)	SA	Decachlorobiphenyl	1.232	1.155	0.998	1.144	1.311	1.168	E7	9.99
3)	L1	AR-1016-1						0.000		-1.00
4)	L1	AR-1016-2						0.000		-1.00
5)	L1	AR-1016-3						0.000		-1.00
6)	L1	AR-1016-4						0.000		-1.00
7)	L1	AR-1016-5						0.000		-1.00
8)	L2	AR-1221-1						0.000		-1.00
9)	L2	AR-1221-2						0.000		-1.00
10)	L2	AR-1221-3						0.000		-1.00
11)	L3	AR-1232-1	3.021	2.873	2.663	3.166	3.542	3.053	E5	10.83
12)	L3	AR-1232-2	1.721	1.670	1.457	1.679	1.815	1.669	E5	7.89
13)	L3	AR-1232-3	3.686	3.539	3.093	3.627	4.127	3.614	E5	10.22
14)	L3	AR-1232-4	1.925	1.863	1.590	1.806	1.973	1.832	E5	8.13
15)	L3	AR-1232-5	1.405	1.349	1.175	1.318	1.403	1.330	E5	7.08
16)	L4	AR-1242-1						0.000		-1.00
17)	L4	AR-1242-2						0.000		-1.00
18)	L4	AR-1242-3						0.000		-1.00
19)	L4	AR-1242-4						0.000		-1.00
20)	L4	AR-1242-5						0.000		-1.00
21)	L5	AR-1248-1						0.000		-1.00
22)	L5	AR-1248-2						0.000		-1.00
23)	L5	AR-1248-3						0.000		-1.00
24)	L5	AR-1248-4						0.000		-1.00
25)	L5	AR-1248-5						0.000		-1.00
26)	L6	AR-1254-1						0.000		-1.00
27)	L6	AR-1254-2						0.000		-1.00
28)	L6	AR-1254-3						0.000		-1.00
29)	L6	AR-1254-4						0.000		-1.00
30)	L6	AR-1254-5						0.000		-1.00
31)	L7	AR-1260-1						0.000		-1.00
32)	L7	AR-1260-2						0.000		-1.00
33)	L7	AR-1260-3						0.000		-1.00
34)	L7	AR-1260-4						0.000		-1.00
35)	L7	AR-1260-5						0.000		-1.00
36)	L8	AR-1262-1						0.000		-1.00
37)	L8	AR-1262-2						0.000		-1.00
38)	L8	AR-1262-3						0.000		-1.00
39)	L8	AR-1262-4						0.000		-1.00
40)	L8	AR-1262-5						0.000		-1.00
41)	L9	AR-1268-1						0.000		-1.00
42)	L9	AR-1268-2						0.000		-1.00
43)	L9	AR-1268-3						0.000		-1.00
44)	L9	AR-1268-4						0.000		-1.00
45)	L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

400 =PR032723.D 1600 =PR032725.D 800 =PR032724.D
 200 =PR032722.D 100 =PR032721.D

Compound		400	1600	800	200	100	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	4.045	4.591	3.996	4.336	4.648	4.323	E7	6.95
2) SA	Decachlorobiphenyl	4.164	4.056	3.416	3.831	4.251	3.944	E7	8.47
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR101018CLP.M
 Title : GC EXTRACTABLES
 Last Update : Thu Oct 11 10:37:15 2018
 Response Via : Initial Calibration

Calibration Files

400 =PR032723.D 1600 =PR032725.D 800 =PR032724.D
 200 =PR032722.D 100 =PR032721.D

	Compound	400	1600	800	200	100	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1	8.398	8.364	7.577	8.754	9.761	8.571 E5	9.25
12) L3	AR-1232-2	1.048	1.088	0.921	1.045	1.176	1.056 E6	8.71
13) L3	AR-1232-3	5.574	5.726	4.737	5.532	5.655	5.444 E5	7.40
14) L3	AR-1232-4	4.791	4.823	4.025	4.692	4.971	4.660 E5	7.92
15) L3	AR-1232-5	5.875	5.552	4.697	5.427	5.772	5.464 E5	8.49
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range