

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR101018CLP.M
 Title : GC EXTRACTABLES
 Last Update : Thu Oct 11 11:28:43 2018
 Response Via : Initial Calibration

Calibration Files

400 =PR032748.D 1600 =PR032750.D 800 =PR032749.D
 200 =PR032747.D 100 =PR032746.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.815	1.692	1.582	1.845	1.951	1.777	E7	8.03
2) SA	Decachlorobiphenyl	2.590	2.499	2.249	2.495	2.790	2.525	E7	7.72
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1						0.000		-1.00
22) L5	AR-1248-2						0.000		-1.00
23) L5	AR-1248-3						0.000		-1.00
24) L5	AR-1248-4						0.000		-1.00
25) L5	AR-1248-5						0.000		-1.00
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1	2.397	2.335	1.927	2.120	2.372	2.230	E6	9.04
42) L9	AR-1268-2	2.239	2.181	1.819	1.980	2.224	2.089	E6	8.75
43) L9	AR-1268-3	2.031	1.997	1.644	1.789	1.992	1.891	E6	8.87
44) L9	AR-1268-4	8.071	7.558	6.775	7.831	8.891	7.825	E5	9.84
45) L9	AR-1268-5	6.014	5.891	5.248	5.699	6.195	5.809	E6	6.24

Signal #2 Calibration Files

400 =PR032748.D 1600 =PR032750.D 800 =PR032749.D
 200 =PR032747.D 100 =PR032746.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	5.450	5.372	4.903	5.475	5.717	5.383	E7	5.54
2) SA	Decachlorobiphenyl	9.294	9.493	8.187	8.835	9.707	9.103	E7	6.64
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR101018CLP.M
 Title : GC EXTRACTABLES
 Last Update : Thu Oct 11 11:28:43 2018
 Response Via : Initial Calibration

Calibration Files

400 =PR032748.D 1600 =PR032750.D 800 =PR032749.D
 200 =PR032747.D 100 =PR032746.D

	Compound	400	1600	800	200	100	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1	1.149	1.166	0.940	1.014	1.103	1.074 E7	8.91
42) L9	AR-1268-2	1.056	1.085	0.873	0.936	1.022	0.994 E7	8.85
43) L9	AR-1268-3	8.811	8.950	7.206	7.795	8.549	8.262 E6	8.96
44) L9	AR-1268-4	3.220	3.130	2.725	3.218	3.399	3.138 E6	8.00
45) L9	AR-1268-5	2.462	2.143	2.203	2.273	2.423	2.301 E7	6.00

(#) = Out of Range