

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR102219\
 Data File : PR042468.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Oct 2019 17:53
 Operator : SM\AJ
 Sample : K5463-12MSD
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 C0AZ0MSD

Manual Integrations
 APPROVED

Ankita
 10/23/2019 1:22:45 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 22 18:26:50 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR100319CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 04 06:59:35 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.724	3.984	88498105	337.1E6	21.755	22.047
2)	SA Decachlor...	10.646	9.092	109.2E6	274.4E6	31.791	27.674
Target Compounds							
3)	L1 AR-1016-1	5.902	5.081	78365177	264.5E6	545.836	486.567
4)	L1 AR-1016-2	5.925	5.101	129.6E6	397.8E6	633.701	551.185
5)	L1 AR-1016-3	5.988	5.279	58092487	190.4E6	477.838	522.053
6)	L1 AR-1016-4	6.088	5.319	60002262	1194.0E6	589.987	4302.659 #
7)	L1 AR-1016-5	6.382	5.535	302.7E6	691.4E6	3040.419	2236.117 #
26)	L6 AR-1254-1	6.759	5.889	944.5E6	2513.4E6	5783.468m	4744.561
27)	L6 AR-1254-2	6.976	6.036	1473.8E6	2553.3E6	5940.459	5092.826
28)	L6 AR-1254-3	7.347	6.443	1686.1E6	4913.4E6	6248.171	5207.939
29)	L6 AR-1254-4	7.632	6.670	1306.0E6	3580.5E6	6434.854	5407.959
30)	L6 AR-1254-5	8.052	7.091	1461.4E6	5154.0E6	6872.105m	6002.474
31)	L7 AR-1260-1	7.509	6.572	664.1E6	2376.6E6	3504.998	4009.402m
32)	L7 AR-1260-2	7.764	6.760	838.9E6	2475.9E6	3549.327	3036.870
33)	L7 AR-1260-3	8.121	6.915	200.6E6	2137.2E6	1068.361	3091.546 #
34)	L7 AR-1260-4	8.356	7.389	638.4E6	376.1E6	3176.528	648.093 #
35)	L7 AR-1260-5	8.703	7.628	372.6E6	1132.0E6	860.805	696.241

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR102219\
Data File : PR042468.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Oct 2019 17:53
Operator : SM\AJ
Sample : K5463-12MSD
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
C0A20MSD

Manual Integrations
APPROVED

Ankita
10/23/2019 1:22:45 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 22 18:26:50 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR100319CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 04 06:59:35 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

