Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR102519\
Data File : PR042565.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Oct 2019 23:15
Operator : SM\AJ
Sample : K5540-15
Misc :
ALS Vial : 49 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
158-88-(0-1)

Manual Integrations APPROVED
Ankita
10/28/2019 10:38:58 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 26 06:02:29 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR101719.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Oct 17 06:40:58 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.725	3.986	70051861	270.2E6	14.709	15.873
2)	SA Decachlor...	10.642	9.086	52952832	153.3E6	15.145	15.106
	Target Compounds						
31)	L7 AR-1260-1	7.507	6.573	9911547	36629892	47.649m	63.329 #
32)	L7 AR-1260-2	7.761	6.757	14385175	39041636	58.585	49.457
33)	L7 AR-1260-3	8.122	6.914	11628883	39722698	62.325	58.409
34)	L7 AR-1260-4	8.358	7.386	16077281	32349486	76.320m	59.040
35)	L7 AR-1260-5	8.700	7.625	25200593	82392555	59.981	55.916

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR102519\
Data File : PR042565.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Oct 2019 23:15
Operator : SM\AJ
Sample : K5540-15
Misc :
ALS Vial : 49 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
158-88-(0-1)

Manual Integrations APPROVED
Ankita
10/28/2019 10:38:58 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 26 06:02:29 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR101719.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Oct 17 06:40:58 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm