

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR102519\
 Data File : PR042543.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Oct 2019 17:14
 Operator : SM\AJ
 Sample : K5503-17
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 171-64-(0-1.8)

Manual Integrations
 APPROVED

Ankita
 10/28/2019 10:37:20 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 26 05:54:16 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR101719.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Oct 17 06:40:58 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.724	3.986	75433943	292.9E6	15.840	17.202
2)	SA Decachlor...	10.641	9.089	52596769	149.8E6	15.043m	14.755
Target Compounds							
31)	L7 AR-1260-1	7.506	6.575	3503543	7268623	16.843	12.567m#
32)	L7 AR-1260-2	7.761	6.758	4433020	10765234	18.054	13.637m
33)	L7 AR-1260-3	8.123	6.915	2068176	8270056	11.084	12.160
34)	L7 AR-1260-4	8.360	7.387	4019394	7125738	19.080	13.005 #
35)	L7 AR-1260-5	8.701	7.626	3257970	9938964	7.754	6.745m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR102519\
Data File : PR042543.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Oct 2019 17:14
Operator : SM\AJ
Sample : K5503-17
Misc :
ALS Vial : 29 Sample Multiplier: 1



Instrument :
ECD_R
ClientSampleId :
171-64-(0-1.8)


Manual Integrations
APPROVED

Ankita
10/28/2019 10:37:20 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 26 05:54:16 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR101719.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Oct 17 06:40:58 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.724 min
Delta R.T.: 0.000 min
Response: 75433943
Conc: 15.84 ng/ml

Instrument :
ECD_R
ClientSampleId :
171-64-(0-1.8)


Manual Integrations
APPROVED

Ankita
10/28/2019 10:37:20 AM




#1 Tetrachloro-m-xylene

R.T.: 3.986 min
Delta R.T.: 0.002 min
Response: 292878905
Conc: 17.20 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.641 min
Delta R.T.: -0.007 min
Response: 52596769
Conc: 15.04 ng/ml m



#2 Decachlorobiphenyl

R.T.: 9.089 min
Delta R.T.: -0.009 min
Response: 149771426
Conc: 14.75 ng/ml




#31 AR-1260-1

R.T.: 7.506 min
Delta R.T.: -0.004 min
Response: 3503543
Conc: 16.84 ng/ml

Instrument :
ECD_R
ClientSampleId :
171-64-(0-1.8)


Manual Integrations
APPROVED

Ankita
10/28/2019 10:37:20 AM




#31 AR-1260-1

R.T.: 6.575 min
Delta R.T.: 0.000 min
Response: 7268623
Conc: 12.57 ng/ml m




#32 AR-1260-2

R.T.: 7.761 min
Delta R.T.: -0.005 min
Response: 4433020
Conc: 18.05 ng/ml



#32 AR-1260-2

R.T.: 6.758 min
Delta R.T.: -0.005 min
Response: 10765234
Conc: 13.64 ng/ml m




#33 AR-1260-3

R.T.: 8.123 min
Delta R.T.: -0.004 min
Response: 2068176
Conc: 11.08 ng/ml

Instrument :
ECD_R
ClientSampleId :
171-64-(0-1.8)


Manual Integrations
APPROVED

Ankita
10/28/2019 10:37:20 AM




#33 AR-1260-3

R.T.: 6.915 min
Delta R.T.: -0.004 min
Response: 8270056
Conc: 12.16 ng/ml




#34 AR-1260-4

R.T.: 8.360 min
Delta R.T.: -0.005 min
Response: 4019394
Conc: 19.08 ng/ml



#34 AR-1260-4

R.T.: 7.387 min
Delta R.T.: -0.007 min
Response: 7125738
Conc: 13.01 ng/ml




#35 AR-1260-5

R.T.: 8.701 min
Delta R.T.: -0.004 min
Response: 3257970
Conc: 7.75 ng/ml

Instrument :
ECD_R
ClientSampleId :
171-64-(0-1.8)

Manual Integrations
APPROVED

Ankita
10/28/2019 10:37:20 AM



#35 AR-1260-5

R.T.: 7.626 min
Delta R.T.: -0.006 min
Response: 9938964
Conc: 6.75 ng/ml m