Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR102819\
Data File : PR042588.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Oct 2019 12:44
Operator : SM\AJ
Sample : K5539-15
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
164-68-(0-1)

Manual Integrations APPROVED
Ankita
10/29/2019 1:30:02 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 29 09:19:53 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR101719.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Oct 17 06:40:58 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.726	3.988	81781025	328.2E6	17.172m	19.276
2)	SA Decachlor...	10.642	9.088	113.2E6	295.6E6	32.366	29.116
	Target Compounds						
31)	L7 AR-1260-1	7.508	6.573	8943969	14465582	42.998m	25.009 #
32)	L7 AR-1260-2	7.762	6.742	4761159	58930840	19.390	74.653m#
33)	L7 AR-1260-3	8.124	6.916	3037444	10765501	16.279m	15.830m
34)	L7 AR-1260-4	8.360	7.387	2659576	12111577	12.625m	22.105 #
35)	L7 AR-1260-5	8.701	7.626	6574539	15442391	15.648	10.480m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR102819\
Data File : PR042588.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Oct 2019 12:44
Operator : SM\AJ
Sample : K5539-15
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
164-68-(0-1)

Manual Integrations APPROVED
Ankita
10/29/2019 1:30:02 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 29 09:19:53 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR101719.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Oct 17 06:40:58 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm