

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR103018\
 Data File : PR033418.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Oct 2018 19:30
 Operator : SM\SJ
 Sample : J5164-09
 Misc : AR1254 MDL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 MDL-WATER-03-QT4-2018

Manual Integrations
 APPROVED

Sohil
 10/31/2018 3:14:57 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 31 08:29:56 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR102818.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Oct 29 02:03:30 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.759	3.740	356.3E6	1307.1E6	20.546	19.994
2) SA Decachlor...	10.750	8.730	302.8E6	721.0E6	18.958	18.954
Target Compounds						
26) L6 AR-1254-1	6.802	5.621	14121512	65303696	17.218m	16.960m
27) L6 AR-1254-2	7.023	5.763	23457524	117.8E6	19.145	36.131m#
28) L6 AR-1254-3	7.395	6.166	22280264	97440872	17.347m	18.852m
29) L6 AR-1254-4	7.681	6.390	17489615	65465988	18.890	21.021m
30) L6 AR-1254-5	8.104	6.805	19944043	113.4E6	16.976m	29.509m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR103018\
Data File : PR033418.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Oct 2018 19:30
Operator : SM\SJ
Sample : J5164-09
Misc : AR1254 MDL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
MDL-WATER-03-QT4-2018

Manual Integrations
APPROVED

Sohil
10/31/2018 3:14:57 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 31 08:29:56 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR102818.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Oct 29 02:03:30 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

