

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR120418\
 Data File : PR034231.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Dec 2018 20:57
 Operator : SM\SJ
 Sample : J6096-08
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 04 04:44:08 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR113018CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Dec 01 01:06:06 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.435	3.524	32016527	63015905	21.891	21.376
2) SA Decachlor...	10.122	8.423	53852920	119.4E6	41.732	41.335
Target Compounds						
7) L1 AR-1016-5	0.000	5.030	0	476447	N.D.	5.170 #
9) L2 AR-1221-2	0.000	3.819	0	686610	N.D.	30.516 #
15) L3 AR-1232-5	0.000	5.030	0	476447	N.D.	11.877 #
24) L5 AR-1248-4	0.000	5.030	0	476447	N.D.	3.309 #
32) L7 AR-1260-2	0.000	6.216	0	1169482	N.D.	4.802 #
45) L9 AR-1268-5	0.000	8.188	0	813582	N.D.	0.513 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR120418\
Data File : PR034231.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Dec 2018 20:57
Operator : SM\SJ
Sample : J6096-08
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 04 04:44:08 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR113018CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Dec 01 01:06:06 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#7 AR-1016-5


R.T.: 0.000 min
Exp R.T.: 6.075 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :




#7 AR-1016-5

R.T.: 5.030 min
Delta R.T.: 0.002 min
Response: 476447
Conc: 5.17 ng/ml




#9 AR-1221-2

R.T.: 0.000 min
Exp R.T.: 4.730 min
Response: 0
Conc: N.D.



#9 AR-1221-2


R.T.: 3.819 min
Delta R.T.: 0.001 min
Response: 686610
Conc: 30.52 ng/ml



#15 AR-1232-5


R.T.: 0.000 min
Exp R.T.: 5.873 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :




#15 AR-1232-5

R.T.: 5.030 min
Delta R.T.: 0.003 min
Response: 476447
Conc: 11.88 ng/ml




#24 AR-1248-4

R.T.: 0.000 min
Exp R.T.: 6.476 min
Response: 0
Conc: N.D.



#24 AR-1248-4


R.T.: 5.030 min
Delta R.T.: 0.003 min
Response: 476447
Conc: 3.31 ng/ml



#32 AR-1260-2


R.T.: 0.000 min
Exp R.T.: 7.449 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :




#32 AR-1260-2

R.T.: 6.216 min
Delta R.T.: -0.012 min
Response: 1169482
Conc: 4.80 ng/ml



#45 AR-1268-5

R.T.: 0.000 min
Exp R.T.: 9.795 min
Response: 0
Conc: N.D.



#45 AR-1268-5

R.T.: 8.188 min
Delta R.T.: -0.001 min
Response: 813582
Conc: 0.51 ng/ml