Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR121118\
Data File : PR034474.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Dec 2018 14:02
Operator : SM\SJ
Sample : J6162-15
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
A0C14-02B2-(0-6)

Manual Integrations APPROVED
Sohil
12/12/2018 2:49:45 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 12 05:32:51 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR120718.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Dec 07 02:02:36 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.431	3.518	31809881	63687749	17.622	17.298
2)	SA Decachlor...	10.122	8.422	31001557	59108160	15.795	14.447
	Target Compounds						
26)	L6 AR-1254-1	6.445	5.367	75666799	225.2E6	929.782m	943.286
27)	L6 AR-1254-2	6.663	5.512	116.8E6	172.2E6	921.983	831.416
28)	L6 AR-1254-3	7.029	5.908	105.3E6	296.8E6	763.292	821.502
29)	L6 AR-1254-4	7.312	6.133	88757341	182.7E6	831.845	788.952
30)	L6 AR-1254-5	7.729	6.546	101.7E6	294.6E6	889.785	910.674

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR121118\
Data File : PR034474.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Dec 2018 14:02
Operator : SM\SJ
Sample : J6162-15
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
A0C14-02B2-(0-6)

Manual Integrations APPROVED
Sohil
12/12/2018 2:49:45 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 12 05:32:51 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR120718.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Dec 07 02:02:36 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm