Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR121219\
Data File : PR043733.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Dec 2019 11:47
Operator : SM\AJ
Sample : K6171-07
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
C0BS9

Manual Integrations APPROVED
Ankita
12/13/2019 11:49:48 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 13 08:29:54 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR120419CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Dec 07 03:51:56 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.704	3.963	101.6E6	133.8E6	18.284	17.324
2)	SA Decachlor...	10.610	9.053	86215817	194.9E6	22.897	19.693
	Target Compounds						
31)	L7 AR-1260-1	7.487	6.548	40484059	74381207	162.463	109.789m#
32)	L7 AR-1260-2	7.742	6.733	45242779	130.7E6	152.283	134.908m
33)	L7 AR-1260-3	8.103	6.889	30218135	95599310	136.398	122.416m
34)	L7 AR-1260-4	8.342	7.360	36682896	67389285	145.835	106.222 #
35)	L7 AR-1260-5	8.679	7.601	58742564	175.5E6	121.817	103.128

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR121219\
Data File : PR043733.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Dec 2019 11:47
Operator : SM\AJ
Sample : K6171-07
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
C0BS9

Manual Integrations APPROVED
Ankita
12/13/2019 11:49:48 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 13 08:29:54 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR120419CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Dec 07 03:51:56 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm