

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR121819\
Data File : PR044016.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Dec 2019 21:22
Operator : SM\AJ
Sample : AR1254CCC400
Misc :
ALS Vial : 6 Sample Multiplier: 1


Instrument :
ECD_R
LabSampleId :
AR1254375

Manual Integrations
APPROVED

Ankita
12/19/2019 10:44:54 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 19 01:12:55 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR121519CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Dec 16 06:15:17 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(26) AR-1254-1 (L6)
R.T. Response Conc
6.75 69512425 363.73
6.96 108635743 356.34
7.33 118689137 344.18
7.62 89025777 339.82
8.04 107939346 380.90

(26) AR-1254-1 #2 (L6)
R.T. Response Conc
5.87 206884552 343.59
6.02 183396812 337.95
6.42 305978406 320.93
6.65 223316490 336.69
7.07 283949752 326.30

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR121819\
Data File : PR044016.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Dec 2019 21:22
Operator : SM\AJ
Sample : AR1254CCC400
Misc :
ALS Vial : 6 Sample Multiplier: 1


Instrument :
ECD_R
LabSampleId :
AR1254375

Manual Integrations
APPROVED

Ankita
12/19/2019 10:44:54 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 19 01:12:55 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR121519CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Dec 16 06:15:17 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(26) AR-1254-1 (L6)
R.T. Response Conc
6.75 69512425 363.73
6.96 108635743 356.34
7.33 118689137 344.18
7.62 89025777 339.82
8.04 94352237 332.95

(26) AR-1254-1 #2 (L6)
R.T. Response Conc
5.87 206884552 343.59
6.02 183396812 337.95
6.42 305978406 320.93
6.65 223316490 336.69
7.07 283949752 326.30

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR121819\
 Data File : PR044016.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Dec 2019 21:22
 Operator : SM\AJ
 Sample : AR1254CCC400
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_R
 LabSampleId :
 AR1254375

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 19 01:12:55 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR121519CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Dec 16 06:15:17 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Manual Integrations
 APPROVED

Ankita
 12/19/2019 10:44:54 AM

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.713	3.971	81284789	217.2E6	20.337	19.677
2) SA Decachlor...	10.620	9.057	164.9E6	361.5E6	34.566	35.837
Target Compounds						
26) L6 AR-1254-1	6.745	5.869	69512425	206.9E6	363.735	343.586
27) L6 AR-1254-2	6.962	6.016	108.6E6	183.4E6	356.335	337.955
28) L6 AR-1254-3	7.332	6.421	118.7E6	306.0E6	344.177	320.932
29) L6 AR-1254-4	7.617	6.649	89025777	223.3E6	339.819	336.687
30) L6 AR-1254-5	8.036	7.067	94352237	283.9E6	332.951m	326.303

3 AJ
 12/20/19

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.