

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR121919\
Data File : PR044039.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Dec 2019 16:00
Operator : SM\AJ
Sample : AR1254CCC400
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_R
LabSampleId :
AR1254377


Manual Integrations
APPROVED

Ankita

12/20/2019 10:47:36 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 19 16:21:04 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR121519CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Dec 16 06:15:17 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(26) AR-1254-1 (L6)
R.T. Response Conc
6.75 78764418 412.15
6.96 127457573 418.07
7.33 142095436 412.05
7.62 108608013 414.57
8.04 133226091 470.13

(26) AR-1254-1 #2 (L6)
R.T. Response Conc
5.87 230020293 382.01
6.02 208778995 384.73
6.42 358302106 375.81
6.65 266357138 401.58
7.07 338548721 389.05

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR121919\
 Data File : PR044039.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Dec 2019 16:00
 Operator : SM\AJ
 Sample : AR1254CCC400
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_R
 LabSampleId :
 AR1254377


Manual Integrations
 APPROVED

Ankita

12/20/2019 10:47:36 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 19 16:21:04 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR121519CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Dec 16 06:15:17 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(26) AR-1254-1 #2 (L6)
 R.T. Response Conc
 6.75 78764418 412.15
 6.96 127457573 418.07
 7.33 142095436 412.05
 7.62 108608013 414.57
 8.04 118586816 418.47

(26) AR-1254-1 #2 (L6)
 R.T. Response Conc
 5.87 230020293 382.01
 6.02 208778995 384.73
 6.42 358302106 375.81
 6.65 266357138 401.58
 7.07 338548721 389.05

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR121919\
 Data File : PR044039.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Dec 2019 16:00
 Operator : SM\AJ
 Sample : AR1254CCC400
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_R
 LabSampleId :
 AR1254377

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 19 16:21:04 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR121519CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Dec 16 06:15:17 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Manual Integrations
 APPROVED

Ankita

12/20/2019 10:47:36 AM

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.714	3.973	83188651	215.7E6	20.814	19.541
2) SA Decachlor...	10.619	9.059	218.1E6	473.0E6	45.707	46.890
Target Compounds						
26) L6 AR-1254-1	6.746	5.869	78764418	230.0E6	412.147	382.009
27) L6 AR-1254-2	6.963	6.015	127.5E6	208.8E6	418.073	384.728
28) L6 AR-1254-3	7.333	6.421	142.1E6	358.3E6	412.051	375.812
29) L6 AR-1254-4	7.617	6.648	108.6E6	266.4E6	414.566	401.579
30) L6 AR-1254-5	8.037	7.068	118.6E6	338.5E6	418.470m	389.046

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

3 AJ
 12/23/19