

Method Path : P:\HPCHEM1\ECD_E\Method\
 Method File : PE100117.M
 Title : 8080.M
 Last Update : Wed Oct 04 13:20:47 2017
 Response Via : Initial Calibration

Calibration Files

750 =PE016828.D 1000 =PE016829.D 1500 =PE016830.D
 500 =PE016827.D 200 =PE016826.D

	Compound	750	1000	1500	500	200	Avg		%RSD
1) T	Dalapon	2.455	2.343	2.274	2.296	2.456	2.365	E2	3.65
2) T	3,5-DICHLOROBENZOI...	1.706	1.534	1.440	1.547	1.622	1.570	E2	6.37
3) T	4-Nitrophenol	9.796	7.978	7.792	7.643	7.814	8.205	E1	10.94
4) S	2,4-DCAA	1.086	0.930	0.925	0.959	0.905	0.961	E2	7.56
5) T	DICAMBA	4.706	4.119	3.933	4.219	4.523	4.300	E2	7.25
6) T	MCP						0.000		-1.00
7) T	MCPA						0.000		-1.00
8) T	DICHLORPROP	1.326	1.186	1.113	1.109	1.209	1.189	E2	7.46
9) T	2,4-D	1.797	1.553	1.516	1.575	1.654	1.619	E2	6.89
10) T	Pentachlorophenol	1.580	1.423	1.362	1.367	1.412	1.429	E3	6.19
11) T	2,4,5-TP (SILVEX)	6.448	5.519	5.496	5.329	5.624	5.683	E2	7.75
12) T	2,4,5-T	7.136	6.254	6.015	6.275	6.512	6.438	E2	6.65
13) T	2,4-DB	1.013	0.908	0.823	0.806	0.962	0.903	E2	9.84
14) T	DINOSEB	5.957	5.203	4.917	5.251	5.519	5.369	E2	7.30
15) T	Picloram	1.187	1.050	1.041	1.037	1.076	1.078	E3	5.82
16) T	DCPA	9.364	8.350	7.904	8.189	9.203	8.602	E2	7.50

Signal #2 Calibration Files

750 =PE016828.D 1000 =PE016829.D 1500 =PE016830.D
 500 =PE016827.D 200 =PE016826.D

	Compound	750	1000	1500	500	200	Avg		%RSD
1) T	Dalapon	3.299	3.140	3.102	3.151	3.474	3.233	E2	4.77
2) T	3,5-DICHLOROBENZOI...	2.905	2.552	2.399	2.643	2.892	2.678	E2	8.18
3) T	4-Nitrophenol	1.638	1.458	1.391	1.413	1.468	1.473	E2	6.59
4) S	2,4-DCAA	1.669	1.418	1.364	1.512	1.584	1.509	E2	8.15
5) T	DICAMBA	7.026	6.040	5.912	6.527	6.930	6.487	E2	7.78
6) T	MCP						0.000		-1.00
7) T	MCPA						0.000		-1.00
8) T	DICHLORPROP	2.302	1.992	1.895	2.055	2.227	2.094	E2	8.01
9) T	2,4-D	2.718	2.265	2.247	2.422	2.468	2.424	E2	7.86
10) T	Pentachlorophenol	2.553	2.257	2.134	2.275	2.387	2.321	E3	6.80
11) T	2,4,5-TP (SILVEX)	1.095	0.965	0.933	0.955	0.977	0.985	E3	6.44
12) T	2,4,5-T	1.198	1.052	0.990	1.051	1.123	1.083	E3	7.36
13) T	2,4-DB	1.407	1.255	1.194	1.245	1.376	1.296	E2	7.05
14) T	DINOSEB	9.669	8.454	7.914	8.664	9.591	8.858	E2	8.53
15) T	Picloram	2.146	1.882	1.819	1.819	1.884	1.910	E3	7.12
16) T	DCPA	1.814	1.599	1.489	1.632	1.759	1.659	E3	7.80

(#) = Out of Range