

Data Path : P:\HPCHEM1\ECD_E\Data\PE100217\
Data File : PE016885.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 3 Oct 2017 18:33
Operator : UA/AJ
Sample : I5546-03
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_E
ClientSampleId :
PAINT-CHIPS-DRUMS-COMP

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 04 16:25:33 2017
Quant Method : P:\HPCHEM1\ECD_E\Method\PE100117.M
Quant Title : 8080.M
QLast Update : Wed Oct 04 13:20:47 2017
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-35-HT Inferno Signal #2 Phase: ZB-XLB-HT Infernon
Signal #1 Info : 30M x 0.25mm x0.2 Signal #2 Info : 30M x 0.25mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

Target Compounds						
10) T	Pentachlo...	12.560	0.000	94107	0	65.864 N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : P:\HPCHEM1\ECD_E\Data\PE100217\
Data File : PE016885.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 3 Oct 2017 18:33
Operator : UA/AJ
Sample : I5546-03
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_E
ClientSampleId :
PAINT-CHIPS-DRUMS-COMP

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 04 16:25:33 2017
Quant Method : P:\HPCHEM1\ECD_E\Method\PE100117.M
Quant Title : 8080.M
QLast Update : Wed Oct 04 13:20:47 2017
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 μ l
Signal #1 Phase : ZB-35-HT Inferno Signal #2 Phase: ZB-XLB-HT Infernon
Signal #1 Info : 30M x 0.25mm x0.2 Signal #2 Info : 30M x 0.25mm x 0.25 μ m






#4 2,4-DCAA


R.T.: 0.000 min
Exp R.T.: 10.110 min
Response: 0
Conc: N.D.

Instrument :
ECD_E
ClientSampleId :
PAINT-CHIPS-DRUMS-COMP




#4 2,4-DCAA

R.T.: 0.000 min
Exp R.T.: 10.696 min
Response: 0
Conc: N.D.



#10 Pentachlorophenol

R.T.: 12.560 min
Delta R.T.: -0.069 min
Response: 94107
Conc: 65.86 ng/ml



#10 Pentachlorophenol

R.T.: 0.000 min
Exp R.T.: 13.032 min
Response: 0
Conc: N.D.