

Method Path : P:\HPCHEM1\ECD_E\Method\
 Method File : PE101217.M
 Title : 8080.M
 Last Update : Fri Oct 13 07:26:24 2017
 Response Via : Initial Calibration

Calibration Files

750 =PE017129.D 1000 =PE017130.D 1500 =PE017131.D
 500 =PE017128.D 200 =PE017127.D

	Compound	750	1000	1500	500	200	Avg		%RSD
1) T	Dalapon	2.096	1.993	1.920	1.994	2.092	2.019	E2	3.69
2) T	3,5-DICHLOROBENZOI...	1.496	1.440	1.373	1.462	1.610	1.476	E2	5.91
3) T	4-Nitrophenol	6.328	5.971	6.105	5.876	5.784	6.013	E1	3.54
4) S	2,4-DCAA	9.967	9.188	8.843	9.662	9.432	9.419	E1	4.58
5) T	DICAMBA	3.680	3.430	3.258	3.569	3.639	3.515	E2	4.91
6) T	MCPP						0.000		-1.00
7) T	MCPA						0.000		-1.00
8) T	DICHLORPROP	1.009	0.948	0.901	0.933	0.940	0.946	E2	4.17
9) T	2,4-D	1.257	1.189	1.119	1.152	1.200	1.183	E2	4.40
10) T	Pentachlorophenol	1.382	1.345	1.293	1.298	1.338	1.331	E3	2.75
11) T	2,4,5-TP (SILVEX)	5.987	5.702	5.191	5.973	6.001	5.771	E2	6.01
12) T	2,4,5-T	5.001	4.670	4.480	4.616	4.832	4.720	E2	4.27
13) T	2,4-DB	7.693	7.100	7.070	7.581	7.102	7.309	E1	4.13
14) T	DINOSEB	4.455	4.137	3.952	4.168	4.306	4.204	E2	4.50
15) T	Picloram	5.581	5.301	5.226	5.174	5.353	5.327	E2	2.96
16) T	DCPA	7.270	6.822	6.534	6.834	7.395	6.971	E2	5.08

Signal #2 Calibration Files

750 =PE017129.D 1000 =PE017130.D 1500 =PE017131.D
 500 =PE017128.D 200 =PE017127.D

	Compound	750	1000	1500	500	200	Avg		%RSD
1) T	Dalapon	2.660	2.514	2.400	2.653	2.911	2.628	E2	7.29
2) T	3,5-DICHLOROBENZOI...	2.388	2.253	2.118	2.350	2.598	2.341	E2	7.58
3) T	4-Nitrophenol	1.075	1.042	1.043	1.005	1.047	1.042	E2	2.37
4) S	2,4-DCAA	1.490	1.398	1.328	1.448	1.566	1.446	E2	6.25
5) T	DICAMBA	5.213	4.910	4.642	5.025	5.657	5.089	E2	7.44
6) T	MCPP						0.000		-1.00
7) T	MCPA						0.000		-1.00
8) T	DICHLORPROP	1.753	1.659	1.548	1.600	1.759	1.664	E2	5.59
9) T	2,4-D	2.128	1.977	1.876	1.779	2.282	2.008	E2	9.98
10) T	Pentachlorophenol	2.292	2.151	2.051	2.150	2.337	2.196	E3	5.30
11) T	2,4,5-TP (SILVEX)	8.772	8.208	7.872	7.915	8.353	8.224	E2	4.46
12) T	2,4,5-T	9.580	8.960	8.518	8.960	9.399	9.083	E2	4.59
13) T	2,4-DB	1.133	1.034	0.980	1.053	1.088	1.057	E2	5.44
14) T	DINOSEB	7.797	7.121	6.667	7.281	7.976	7.369	E2	7.16
15) T	Picloram	1.170	1.125	1.140	1.177	1.059	1.134	E3	4.15
16) T	DCPA	1.484	1.389	1.334	1.387	1.627	1.444	E3	8.01

(#) = Out of Range