

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_A\Data\FA033122\
Data File : FA003305.D
Signal(s) : FID1A.CH
Acq On : 31 Mar 2022 10:00
Operator : SJ/MA
Sample : 500 PPMV GASES ICC
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
FID_A
ClientSampleId :

Integration File: autoint1.e
Quant Time: Mar 31 09:56:56 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_A\Method\FA033122 GASES.M
Quant Title :
QLast Update : Thu Mar 31 09:55:27 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. :
Signal Phase : RT-U Plot
Signal Info : 30M x 0.32mm

Compound	R.T.	Response	Conc Units

Target Compounds			
1) Ethane	1.796	101612	421.626 ppmv
2) Ethylene	1.708	95113	412.377 ppmv
3) Methane	1.323	54180	406.002 ppmv

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_A\Data\FA033122\
Data File : FA003305.D
Signal(s) : FID1A.CH
Acq On : 31 Mar 2022 10:00
Operator : SJ/MA
Sample : 500 PPMV GASES ICC
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
FID_A
ClientSampleId :

Integration File: autoint1.e
Quant Time: Mar 31 09:56:56 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_A\Method\FA033122 GASES.M
Quant Title :
QLast Update : Thu Mar 31 09:55:27 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. :
Signal Phase : RT-U Plot
Signal Info : 30M x 0.32mm






#1 Ethane

R.T.: 1.796 min
Delta R.T.: -0.006 min
Response: 101612
Conc: 421.63 ppmv

Instrument :


FID_A

ClientSampleId :



#2 Ethylene

R.T.: 1.708 min
Delta R.T.: -0.006 min
Response: 95113
Conc: 412.38 ppmv



#3 Methane

R.T.: 1.323 min
Delta R.T.: -0.003 min
Response: 54180
Conc: 406.00 ppmv