

Data Path : Y:\bestpcbsrv\HPCHEM1\FID A\Data\FA060718\
Data File : FA002061.D
Signal(s) : FID1A.CH
Acq On : 7 Jun 2018 19:24
Operator : UA/AJ
Sample : 200 PPMV CCC
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
FID_A
ClientSampleId :
200 PPMV CCC

Integration File: autoint1.e
Quant Time: Jun 07 19:31:32 2018
Quant Method : Y:\pestpcbsrv\HPCHEM1\FID_A\Method\FA060718 GASES.M
Quant Title :
QLast Update : Thu Jun 07 16:13:43 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. :
Signal Phase : RT-U Plot
Signal Info : 30M x 0.32mm

	Compound	R.T.	Response	Conc	Units

Target Compounds					
1)	Ethane	3.211	25397	189.944	ppmv
2)	Ethylene	2.974	21229	174.804	ppmv
3)	Methane	2.090	15638	179.619	ppmv

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Y:\pestpcbsrv\HPCHEM1\FID A\Data\FA060718\
Data File : FA002061.D
Signal(s) : FID1A.CH
Acq On : 7 Jun 2018 19:24
Operator : UA/AJ
Sample : 200 PPMV CCC
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
FID_A
ClientSampleId :
200 PPMV CCC

Integration File: autoint1.e
Quant Time: Jun 07 19:31:32 2018
Quant Method : Y:\pestpcbsrv\HPCHEM1\FID_A\Method\FA060718 GASES.M
Quant Title :
QLast Update : Thu Jun 07 16:13:43 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. :
Signal Phase : RT-U Plot
Signal Info : 30M x 0.32mm






#1 Ethane


R.T.: 3.211 min
Delta R.T.: -0.014 min
Response: 25397
Conc: 189.94 ppmv

Instrument :
FID_A
ClientSampleId :
200 PPMV CCC



#2 Ethylene

R.T.: 2.974 min
Delta R.T.: -0.013 min
Response: 21229
Conc: 174.80 ppmv



#3 Methane

R.T.: 2.090 min
Delta R.T.: -0.010 min
Response: 15638
Conc: 179.62 ppmv