

Data Path : Z:\bestpcbsrv\HPCHEM1\FID A\Data\FA061718\
Data File : FA002087.D
Signal(s) : FID1A.CH
Acq On : 17 Jun 2018 16:38
Operator : UA/AJ
Sample : 200 PPMV CCC
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
FID_A
ClientSampleId :
200 PPMV CCC

Integration File: autoint1.e
Quant Time: Jun 22 11:00:28 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_A\Method\FA060718 GASES.M
Quant Title :
QLast Update : Fri Jun 22 10:56:52 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. :
Signal Phase : RT-U Plot
Signal Info : 30M x 0.32mm

	Compound	R.T.	Response	Conc	Units

Target Compounds					
1)	Ethane	3.034	25215	188.588	ppmv
2)	Ethylene	2.840	21923	180.518	ppmv
3)	Methane	2.077	16863	193.688	ppmv

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID A\Data\FA061718\
Data File : FA002087.D
Signal(s) : FID1A.CH
Acq On : 17 Jun 2018 16:38
Operator : UA/AJ
Sample : 200 PPMV CCC
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
FID_A
ClientSampleId :
200 PPMV CCC

Integration File: autoint1.e
Quant Time: Jun 22 11:00:28 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_A\Method\FA060718 GASES.M
Quant Title :
QLast Update : Fri Jun 22 10:56:52 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. :
Signal Phase : RT-U Plot
Signal Info : 30M x 0.32mm






#1 Ethane


R.T.: 3.034 min
Delta R.T.: 0.004 min
Response: 25215
Conc: 188.59 ppmv

Instrument :
FID_A
ClientSampleId :
200 PPMV CCC



#2 Ethylene

R.T.: 2.840 min
Delta R.T.: 0.000 min
Response: 21923
Conc: 180.52 ppmv



#3 Methane

R.T.: 2.077 min
Delta R.T.: 0.007 min
Response: 16863
Conc: 193.69 ppmv