

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_A\Data\FA091222\
Data File : FA003467.D
Signal(s) : FID1A.CH
Acq On : 12 Sep 2022 11:15
Operator : SJ/MA
Sample : 200 PPMV GASES CCC
Misc :
ALS Vial : 7 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Sep 12 11:11:36 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_A\Method\FA091222 GASES.M
Quant Title :
QLast Update : Mon Sep 12 10:58:21 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. :
Signal Phase : RT-U Plot
Signal Info : 30M x 0.32mm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1	Ethane	193.141	175.776	9.0	92	0.00
2	Ethylene	204.807	177.657	13.3	87	0.00
3	Methane	167.216	151.684	9.3	92	0.00

Evaluate Continuing Calibration Report - Not Founds

(#) = Out of Range

SPCC's out = 0 CCC's out = 0