

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_A\Data\FA100322\
Data File : FA003515.D
Signal(s) : FID1A.CH
Acq On : 3 Oct 2022 11:02
Operator : SJ/MA
Sample : 50 PPMV GASES ICC
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
FID_A
ClientSampleId :
50 PPMV GASES ICC

Integration File: autoint1.e
Quant Time: Oct 03 11:01:01 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_A\Method\FA100322 GASES.M
Quant Title :
QLast Update : Mon Oct 03 10:50:25 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. :
Signal Phase : RT-U Plot
Signal Info : 30M x 0.32mm

Compound	R.T.	Response	Conc Units

Target Compounds			
1) Ethane	1.676	4024	48.821 ppmv
2) Ethylene	1.624	4214	47.077 ppmv
3) Methane	1.314	4958	51.173 ppmv

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_A\Data\FA100322\
Data File : FA003515.D
Signal(s) : FID1A.CH
Acq On : 3 Oct 2022 11:02
Operator : SJ/MA
Sample : 50 PPMV GASES ICC
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
FID_A
ClientSampleId :
50 PPMV GASES ICC

Integration File: autoint1.e
Quant Time: Oct 03 11:01:01 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_A\Method\FA100322 GASES.M
Quant Title :
QLast Update : Mon Oct 03 10:50:25 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. :
Signal Phase : RT-U Plot
Signal Info : 30M x 0.32mm






#1 Ethane


R.T.: 1.676 min
Delta R.T.: -0.010 min
Response: 4024
Conc: 48.82 ppmv

Instrument :
FID_A
ClientSampleId :
50 PPMV GASES ICC



#2 Ethylene

R.T.: 1.624 min
Delta R.T.: -0.005 min
Response: 4214
Conc: 47.08 ppmv



#3 Methane

R.T.: 1.314 min
Delta R.T.: -0.002 min
Response: 4958
Conc: 51.17 ppmv