

Method Path : Z:\pestpcbsrv\HPCHEM1\FID_A\Method\
Method File : FA100621 GASES.M
Title :
Last Update : Wed Oct 06 09:56:07 2021
Response Via : Initial Calibration

Calibration Files

500 =FA003199.D 300 =FA003198.D 200 =FA003197.D
100 =FA003196.D 50 =FA003195.D

	Compound	500	300	200	100	50	Avg		%RSD
1)	Ethane	1.608	1.776	1.715	1.782	1.815	1.739	E2	4.70
2)	Ethylene	1.640	1.826	1.806	1.846	1.889	1.802	E2	5.30
3)	Methane	1.217	1.335	1.392	1.450	1.591	1.397	E2	9.91

(#) = Out of Range