

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_A\Data\FA100818\
Data File : FA002290.D
Signal(s) : FID1A.CH
Acq On : 8 Oct 2018 14:53
Operator : UA/AJ
Sample : J5164-07
Misc : LOD 10 PPMV
ALS Vial : 15 Sample Multiplier: 1

Instrument :
FID_A
ClientSampleId :
LOD-MDL-WATER-01-QT4-2018

Integration File: autoint1.e
Quant Time: Oct 09 09:07:35 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_A\Method\FA092718 GASES.M
Quant Title :
QLast Update : Thu Sep 27 09:41:44 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. :
Signal Phase : RT-U Plot
Signal Info : 30M x 0.32mm

Compound	R.T.	Response	Conc Units

Target Compounds			
1) Ethane	2.485	2820	9.683 ppmv
2) Ethylene	2.399	2470	9.398 ppmv
3) Methane	1.994	1994	11.871 ppmv

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_A\Data\FA100818\
Data File : FA002290.D
Signal(s) : FID1A.CH
Acq On : 8 Oct 2018 14:53
Operator : UA/AJ
Sample : J5164-07
Misc : LOD 10 PPMV
ALS Vial : 15 Sample Multiplier: 1

Instrument :
FID_A
ClientSampleId :
LOD-MDL-WATER-01-QT4-2018

Integration File: autoint1.e
Quant Time: Oct 09 09:07:35 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_A\Method\FA092718 GASES.M
Quant Title :
QLast Update : Thu Sep 27 09:41:44 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. :
Signal Phase : RT-U Plot
Signal Info : 30M x 0.32mm






#1 Ethane


R.T.: 2.485 min
Delta R.T.: -0.021 min
Response: 2820
Conc: 9.68 ppmv

Instrument :
FID_A
ClientSampleId :
LOD-MDL-WATER-01-QT4-2018



#2 Ethylene

R.T.: 2.399 min
Delta R.T.: -0.020 min
Response: 2470
Conc: 9.40 ppmv



#3 Methane

R.T.: 1.994 min
Delta R.T.: -0.001 min
Response: 1994
Conc: 11.87 ppmv