

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_A\Data\FA102921\
Data File : FA003230.D
Signal(s) : FID1A.CH
Acq On : 29 Oct 2021 10:35
Operator : SJ/MA
Sample : 200 PPMV GASES CCC
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
FID_A
ClientSampleId :

Integration File: autoint1.e
Quant Time: Oct 29 10:36:50 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_A\Method\FA102921 GASES.M
Quant Title :
QLast Update : Fri Oct 29 09:09:00 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. :
Signal Phase : RT-U Plot
Signal Info : 30M x 0.32mm

Compound	R.T.	Response	Conc Units

Target Compounds			
1) Ethane	1.800	39978	210.771 ppmv
2) Ethylene	1.710	35557	186.880 ppmv
3) Methane	1.319	29572	217.711 ppmv

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_A\Data\FA102921\
Data File : FA003230.D
Signal(s) : FID1A.CH
Acq On : 29 Oct 2021 10:35
Operator : SJ/MA
Sample : 200 PPMV GASES CCC
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
FID_A
ClientSampleId :

Integration File: autoint1.e
Quant Time: Oct 29 10:36:50 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_A\Method\FA102921 GASES.M
Quant Title :
QLast Update : Fri Oct 29 09:09:00 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. :
Signal Phase : RT-U Plot
Signal Info : 30M x 0.32mm






#1 Ethane


R.T.: 1.800 min
Delta R.T.: -0.004 min
Response: 39978
Conc: 210.77 ppmv

Instrument :
FID_A
ClientSampleId :



#2 Ethylene

R.T.: 1.710 min
Delta R.T.: -0.003 min
Response: 35557
Conc: 186.88 ppmv



#3 Methane

R.T.: 1.319 min
Delta R.T.: -0.002 min
Response: 29572
Conc: 217.71 ppmv