

Method Path : P:\HPCHEM1\FID E\Method\
Method File : FE022713 GASES.M
Title :
Last Update : Wed Feb 27 18:56:14 2013
Response Via : Initial Calibration

Calibration Files

200 =FE007861.D 1000 =FE007859.D 500 =FE007860.D
100 =FE007862.D 50 =FE007863.D

	Compound	200	1000	500	100	50	Avg		%RSD
1)	Ethane	2.064	1.393	1.616	1.966	1.813	1.770	E4	15.29
2)	Ethylene	2.081	1.405	1.630	1.985	1.823	1.785	E4	15.29
3)	Methane	1.153	0.763	0.871	1.143	1.139	1.014	E4	18.13

(#) = Out of Range