

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB010319\
 Data File : FB019249.D
 Signal(s) : FID2B.CH
 Acq On : 3 Jan 2019 10:39
 Operator : MJ/SJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :

Integration File: Calibration.e
 Quant Time: Jan 03 12:57:09 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB120718.M
 Quant Title :
 QLast Update : Fri Dec 07 17:57:38 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	9.107	170004	19.450 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.961	422241	27.620 ng/ml
2) t 2,2,4-Trimethylpentane	7.724	442595	28.729 ng/ml
3) t n-Heptane	8.063	143057	9.625 ng/ml
4) t Benzene	8.201	178734	9.923 ng/ml
6) t Toluene	10.956	466121	30.708 ng/ml
7) t Ethylbenzene	13.406	160409	10.325 ng/ml
8) t m-Xylene	13.539	338663	20.788 ng/ml
9) t O-Xylene	14.271	355516	21.026 ng/ml
10) t 1,2,4-Trimethylbenzene	16.554	411759	20.346 ng/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB010319\
Data File : FB019249.D
Signal(s) : FID2B.CH
Acq On : 3 Jan 2019 10:39
Operator : MJ/SJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :

Integration File: Calibration.e
Quant Time: Jan 03 12:57:09 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB120718.M
Quant Title :
QLast Update : Fri Dec 07 17:57:38 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane


R.T.: 4.961 min
Delta R.T.: 0.000 min
Response: 422241
Conc: 27.62 ng/ml

Instrument :
FID_B
ClientSampleId :




#2 2,2,4-Trimethylpentane

R.T.: 7.724 min
Delta R.T.: 0.001 min
Response: 442595
Conc: 28.73 ng/ml




#3 n-Heptane

R.T.: 8.063 min
Delta R.T.: 0.000 min
Response: 143057
Conc: 9.62 ng/ml



#4 Benzene


R.T.: 8.201 min
Delta R.T.: 0.000 min
Response: 178734
Conc: 9.92 ng/ml



#5 AAA-TFT


R.T.: 9.107 min
Delta R.T.: 0.000 min
Response: 170004
Conc: 19.45 ng/ml

Instrument :
FID_B
ClientSampleId :




#6 Toluene

R.T.: 10.956 min
Delta R.T.: 0.000 min
Response: 466121
Conc: 30.71 ng/ml




#7 Ethylbenzene

R.T.: 13.406 min
Delta R.T.: 0.000 min
Response: 160409
Conc: 10.33 ng/ml



#8 m-Xylene


R.T.: 13.539 min
Delta R.T.: 0.001 min
Response: 338663
Conc: 20.79 ng/ml



#9 O-Xylene

R.T.: 14.271 min
Delta R.T.: 0.000 min
Response: 355516
Conc: 21.03 ng/ml

Instrument :
FID_B
ClientSampleId :



#10 1,2,4-Trimethylbenzene

R.T.: 16.554 min
Delta R.T.: 0.000 min
Response: 411759
Conc: 20.35 ng/ml