

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB010819\
 Data File : FB019295.D
 Signal(s) : FID2B.CH
 Acq On : 8 Jan 2019 21:25
 Operator : MJ/SJ
 Sample : FB010819GROICV
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Jan 09 07:22:52 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB010819.M
 Quant Title :
 QLast Update : Wed Jan 09 07:19:31 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	9.045	165607	18.637 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.900	553716	27.802 ng/ml
2) t 2,2,4-Trimethylpentane	7.660	538918	28.290 ng/ml
3) t n-Heptane	7.999	163123	9.349 ng/ml
4) t Benzene	8.137	189681	9.660 ng/ml
6) t Toluene	10.891	537657	29.723 ng/ml
7) t Ethylbenzene	13.343	221548	10.000 ng/ml
8) t m-Xylene	13.477	457324	19.928 ng/ml
9) t O-Xylene	14.208	451853	20.138 ng/ml
10) t 1,2,4-Trimethylbenzene	16.495	452129	19.828 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB010819\
Data File : FB019295.D
Signal(s) : FID2B.CH
Acq On : 8 Jan 2019 21:25
Operator : MJ/SJ
Sample : FB010819GROICV
Misc :
ALS Vial : 10 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Jan 09 07:22:52 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB010819.M
Quant Title :
QLast Update : Wed Jan 09 07:19:31 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.900 min
Delta R.T.: 0.003 min
Response: 553716
Conc: 27.80 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.660 min
Delta R.T.: 0.002 min
Response: 538918
Conc: 28.29 ng/ml




#3 n-Heptane

R.T.: 7.999 min
Delta R.T.: 0.000 min
Response: 163123
Conc: 9.35 ng/ml




#4 Benzene

R.T.: 8.137 min
Delta R.T.: 0.001 min
Response: 189681
Conc: 9.66 ng/ml




#5 AAA-TFT

R.T.: 9.045 min
Delta R.T.: 0.002 min
Response: 165607
Conc: 18.64 ng/ml




#6 Toluene

R.T.: 10.891 min
Delta R.T.: 0.001 min
Response: 537657
Conc: 29.72 ng/ml




#7 Ethylbenzene

R.T.: 13.343 min
Delta R.T.: 0.001 min
Response: 221548
Conc: 10.00 ng/ml




#8 m-Xylene

R.T.: 13.477 min
Delta R.T.: 0.000 min
Response: 457324
Conc: 19.93 ng/ml



#9 O-Xylene

R.T.: 14.208 min
Delta R.T.: 0.000 min
Response: 451853
Conc: 20.14 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.495 min
Delta R.T.: 0.001 min
Response: 452129
Conc: 19.83 ng/ml