

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB010824\
 Data File : FB030532.D
 Signal(s) : FID2B.CH
 Acq On : 8 Jan 2024 13:06
 Operator : YP/AJ
 Sample : 100 GRO STD
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Jan 08 12:50:42 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB010824.M
 Quant Title :
 QLast Update : Mon Jan 08 12:48:49 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.780	1361889	96.228 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.698	2306229	143.916 ng/ml
2) t 2,2,4-Trimethylpentane	7.405	2484950	146.370 ng/ml
3) t n-Heptane	7.739	877115	49.677 ng/ml
4) t Benzene	7.875	1151835	50.523 ng/ml
6) t Toluene	10.608	3346772	151.099 ng/ml
7) t Ethylbenzene	13.045	1018401	49.418 ng/ml
8) t m-Xylene	13.178	2005946	98.351 ng/ml
9) t O-Xylene	13.905	1933364	97.583 ng/ml
10) t 1,2,4-Trimethylbenzene	16.180	1811120	93.716 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB010824\
Data File : FB030532.D
Signal(s) : FID2B.CH
Acq On : 8 Jan 2024 13:06
Operator : YP/AJ
Sample : 100 GRO STD
Misc :
ALS Vial : 5 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Jan 08 12:50:42 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB010824.M
Quant Title :
QLast Update : Mon Jan 08 12:48:49 2024
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.698 min
Delta R.T.: 0.000 min
Response: 2306229
Conc: 143.92 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.405 min
Delta R.T.: 0.000 min
Response: 2484950
Conc: 146.37 ng/ml




#3 n-Heptane

R.T.: 7.739 min
Delta R.T.: 0.000 min
Response: 877115
Conc: 49.68 ng/ml




#4 Benzene

R.T.: 7.875 min
Delta R.T.: 0.000 min
Response: 1151835
Conc: 50.52 ng/ml




#5 AAA-TFT

R.T.: 8.780 min
Delta R.T.: 0.000 min
Response: 1361889
Conc: 96.23 ng/ml




#6 Toluene

R.T.: 10.608 min
Delta R.T.: 0.000 min
Response: 3346772
Conc: 151.10 ng/ml




#7 Ethylbenzene

R.T.: 13.045 min
Delta R.T.: 0.000 min
Response: 1018401
Conc: 49.42 ng/ml




#8 m-Xylene

R.T.: 13.178 min
Delta R.T.: 0.000 min
Response: 2005946
Conc: 98.35 ng/ml



#9 O-Xylene

R.T.: 13.905 min
Delta R.T.: 0.000 min
Response: 1933364
Conc: 97.58 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.180 min
Delta R.T.: 0.000 min
Response: 1811120
Conc: 93.72 ng/ml