

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB010919\
 Data File : FB019345.D
 Signal(s) : FID2B.CH
 Acq On : 10 Jan 2019 9:14
 Operator : MJ/SJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Jan 10 09:38:06 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB010819.M
 Quant Title :
 QLast Update : Wed Jan 09 07:19:31 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	9.044	163541	18.404 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.898	548361	27.533 ng/ml
2) t 2,2,4-Trimethylpentane	7.659	538117	28.248 ng/ml
3) t n-Heptane	7.999	163309	9.360 ng/ml
4) t Benzene	8.137	191305	9.743 ng/ml
6) t Toluene	10.892	529708	29.283 ng/ml
7) t Ethylbenzene	13.345	210735	9.512 ng/ml
8) t m-Xylene	13.478	433912	18.908 ng/ml
9) t O-Xylene	14.210	429974	19.163 ng/ml
10) t 1,2,4-Trimethylbenzene	16.496	419192	18.384 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB010919\
Data File : FB019345.D
Signal(s) : FID2B.CH
Acq On : 10 Jan 2019 9:14
Operator : MJ/SJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 35 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Jan 10 09:38:06 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB010819.M
Quant Title :
QLast Update : Wed Jan 09 07:19:31 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.898 min
Delta R.T.: 0.000 min
Response: 548361
Conc: 27.53 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.659 min
Delta R.T.: 0.002 min
Response: 538117
Conc: 28.25 ng/ml




#3 n-Heptane

R.T.: 7.999 min
Delta R.T.: 0.001 min
Response: 163309
Conc: 9.36 ng/ml




#4 Benzene

R.T.: 8.137 min
Delta R.T.: 0.002 min
Response: 191305
Conc: 9.74 ng/ml




#5 AAA-TFT

R.T.: 9.044 min
Delta R.T.: 0.000 min
Response: 163541
Conc: 18.40 ng/ml




#6 Toluene

R.T.: 10.892 min
Delta R.T.: 0.002 min
Response: 529708
Conc: 29.28 ng/ml




#7 Ethylbenzene

R.T.: 13.345 min
Delta R.T.: 0.002 min
Response: 210735
Conc: 9.51 ng/ml




#8 m-Xylene

R.T.: 13.478 min
Delta R.T.: 0.002 min
Response: 433912
Conc: 18.91 ng/ml



#9 O-Xylene

R.T.: 14.210 min
Delta R.T.: 0.002 min
Response: 429974
Conc: 19.16 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.496 min
Delta R.T.: 0.002 min
Response: 419192
Conc: 18.38 ng/ml