

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB010920\
 Data File : FB023393.D
 Signal(s) : FID2B.CH
 Acq On : 9 Jan 2020 17:14
 Operator : DD/AJ
 Sample : L1070-05
 Misc : 5.02g/5.00ml DI WATER
 ALS Vial : 7 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Jan 10 00:18:20 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB121919.M
 Quant Title :
 QLast Update : Thu Dec 19 12:43:06 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.791	496596	20.941 ng/ml
Target Compounds			
2) t 2,2,4-Trimethylpentane	7.442	7272	0.218 ng/ml
3) t n-Heptane	7.745	5798	0.176 ng/ml
7) t Ethylbenzene	13.023	196397	5.411 ng/ml
9) t O-Xylene	13.936	30836	0.915 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB010920\
Data File : FB023393.D
Signal(s) : FID2B.CH
Acq On : 9 Jan 2020 17:14
Operator : DD/AJ
Sample : L1070-05
Misc : 5.02g/5.00ml DI WATER
ALS Vial : 7 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Jan 10 00:18:20 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB121919.M
Quant Title :
QLast Update : Thu Dec 19 12:43:06 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#2 2,2,4-Trimethylpentane

R.T.: 7.442 min
Delta R.T.: 0.030 min
Response: 7272
Conc: 0.22 ng/ml




#3 n-Heptane

R.T.: 7.745 min
Delta R.T.: -0.007 min
Response: 5798
Conc: 0.18 ng/ml




#5 AAA-TFT

R.T.: 8.791 min
Delta R.T.: -0.002 min
Response: 496596
Conc: 20.94 ng/ml



#7 Ethylbenzene

R.T.: 13.023 min
Delta R.T.: -0.041 min
Response: 196397
Conc: 5.41 ng/ml



#9 O-Xylene

R.T.: 13.936 min
Delta R.T.: 0.011 min
Response: 30836
Conc: 0.92 ng/ml