

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB011019\
 Data File : FB019346.D
 Signal(s) : FID2B.CH
 Acq On : 10 Jan 2019 9:44
 Operator : MJ/SJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :

Integration File: Calibration.e
 Quant Time: Jan 11 04:50:42 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB010819.M
 Quant Title :
 QLast Update : Wed Jan 09 07:19:31 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	9.044	167273	18.824 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.898	536979	26.962 ng/ml
2) t 2,2,4-Trimethylpentane	7.659	510965	26.822 ng/ml
3) t n-Heptane	7.999	152987	8.768 ng/ml
4) t Benzene	8.136	180326	9.184 ng/ml
6) t Toluene	10.891	525815	29.068 ng/ml
7) t Ethylbenzene	13.343	207164	9.350 ng/ml
8) t m-Xylene	13.477	429428	18.713 ng/ml
9) t O-Xylene	14.208	422015	18.808 ng/ml
10) t 1,2,4-Trimethylbenzene	16.495	411908	18.064 ng/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB011019\
Data File : FB019346.D
Signal(s) : FID2B.CH
Acq On : 10 Jan 2019 9:44
Operator : MJ/SJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 36 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :

Integration File: Calibration.e
Quant Time: Jan 11 04:50:42 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB010819.M
Quant Title :
QLast Update : Wed Jan 09 07:19:31 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane


R.T.: 4.898 min
Delta R.T.: 0.000 min
Response: 536979
Conc: 26.96 ng/ml

Instrument :
FID_B
ClientSampleId :




#2 2,2,4-Trimethylpentane

R.T.: 7.659 min
Delta R.T.: 0.000 min
Response: 510965
Conc: 26.82 ng/ml




#3 n-Heptane

R.T.: 7.999 min
Delta R.T.: 0.000 min
Response: 152987
Conc: 8.77 ng/ml



#4 Benzene


R.T.: 8.136 min
Delta R.T.: 0.000 min
Response: 180326
Conc: 9.18 ng/ml



#5 AAA-TFT


R.T.: 9.044 min
Delta R.T.: 0.000 min
Response: 167273
Conc: 18.82 ng/ml

Instrument :
FID_B
ClientSampleId :




#6 Toluene

R.T.: 10.891 min
Delta R.T.: 0.001 min
Response: 525815
Conc: 29.07 ng/ml




#7 Ethylbenzene

R.T.: 13.343 min
Delta R.T.: 0.001 min
Response: 207164
Conc: 9.35 ng/ml



#8 m-Xylene


R.T.: 13.477 min
Delta R.T.: 0.001 min
Response: 429428
Conc: 18.71 ng/ml



#9 O-Xylene

R.T.: 14.208 min
Delta R.T.: 0.000 min
Response: 422015
Conc: 18.81 ng/ml

Instrument :
FID_B
ClientSampleId :



#10 1,2,4-Trimethylbenzene

R.T.: 16.495 min
Delta R.T.: 0.001 min
Response: 411908
Conc: 18.06 ng/ml