

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB011022\
 Data File : FB026585.D
 Signal(s) : FID2B.CH
 Acq On : 10 Jan 2022 11:47
 Operator : AJ/MA
 Sample : 20 GRO STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :

Integration File: Calibration.e
 Quant Time: Jan 10 12:33:11 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB011022.M
 Quant Title :
 QLast Update : Mon Jan 10 12:32:56 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.786	204763	20.000 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.704	471190	30.000 ng/ml
2) t 2,2,4-Trimethylpentane	7.425	548989	30.000 ng/ml
3) t n-Heptane	7.745	177362	10.000 ng/ml
4) t Benzene	7.880	203701	10.000 ng/ml
6) t Toluene	10.613	588309	30.000 ng/ml
7) t Ethylbenzene	13.050	184555	10.000 ng/ml
8) t m-Xylene	13.182	368577	20.000 ng/ml
9) t O-Xylene	13.910	355922	20.000 ng/ml
10) t 1,2,4-Trimethylbenzene	16.185	366309	20.000 ng/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB011022\
Data File : FB026585.D
Signal(s) : FID2B.CH
Acq On : 10 Jan 2022 11:47
Operator : AJ/MA
Sample : 20 GRO STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :

Integration File: Calibration.e
Quant Time: Jan 10 12:33:11 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB011022.M
Quant Title :
QLast Update : Mon Jan 10 12:32:56 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane


R.T.: 4.704 min
Delta R.T.: 0.000 min
Response: 471190
Conc: 30.00 ng/ml

Instrument :
FID_B
ClientSampleId :




#2 2,2,4-Trimethylpentane

R.T.: 7.425 min
Delta R.T.: 0.000 min
Response: 548989
Conc: 30.00 ng/ml




#3 n-Heptane

R.T.: 7.745 min
Delta R.T.: 0.000 min
Response: 177362
Conc: 10.00 ng/ml



#4 Benzene


R.T.: 7.880 min
Delta R.T.: 0.000 min
Response: 203701
Conc: 10.00 ng/ml



#5 AAA-TFT


R.T.: 8.786 min
Delta R.T.: 0.000 min
Response: 204763
Conc: 20.00 ng/ml

Instrument :
FID_B
ClientSampleId :




#6 Toluene

R.T.: 10.613 min
Delta R.T.: 0.000 min
Response: 588309
Conc: 30.00 ng/ml




#7 Ethylbenzene

R.T.: 13.050 min
Delta R.T.: 0.000 min
Response: 184555
Conc: 10.00 ng/ml



#8 m-Xylene


R.T.: 13.182 min
Delta R.T.: 0.000 min
Response: 368577
Conc: 20.00 ng/ml



#9 O-Xylene

R.T.: 13.910 min
Delta R.T.: 0.000 min
Response: 355922
Conc: 20.00 ng/ml

Instrument :
FID_B
ClientSampleId :



#10 1,2,4-Trimethylbenzene

R.T.: 16.185 min
Delta R.T.: 0.000 min
Response: 366309
Conc: 20.00 ng/ml