

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB011420\
 Data File : FB023406.D
 Signal(s) : FID2B.CH
 Acq On : 14 Jan 2020 10:35
 Operator : DD/AJ
 Sample : L1004-05
 Misc : 5.02g/5.00ml DI WATER
 ALS Vial : 4 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Jan 14 21:09:11 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB121919.M
 Quant Title :
 QLast Update : Thu Dec 19 12:43:06 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.791	492278	20.759 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.664	611	0.020 ng/ml
2) t 2,2,4-Trimethylpentane	7.400	235	0.007 ng/ml
6) t Toluene	10.619	4875	0.127 ng/ml
8) t m-Xylene	13.204	4264	0.123 ng/ml
9) t O-Xylene	13.916	5124	0.152 ng/ml
10) t 1,2,4-Trimethylbenzene	16.208	10894	0.388 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB011420\
Data File : FB023406.D
Signal(s) : FID2B.CH
Acq On : 14 Jan 2020 10:35
Operator : DD/AJ
Sample : L1004-05
Misc : 5.02g/5.00ml DI WATER
ALS Vial : 4 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Jan 14 21:09:11 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB121919.M
Quant Title :
QLast Update : Thu Dec 19 12:43:06 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.664 min
Delta R.T.: -0.020 min
Response: 611
Conc: 0.02 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.400 min
Delta R.T.: -0.011 min
Response: 235
Conc: 0.01 ng/ml




#5 AAA-TFT

R.T.: 8.791 min
Delta R.T.: -0.002 min
Response: 492278
Conc: 20.76 ng/ml




#6 Toluene

R.T.: 10.619 min
Delta R.T.: -0.005 min
Response: 4875
Conc: 0.13 ng/ml




#8 m-Xylene

R.T.: 13.204 min
Delta R.T.: 0.007 min
Response: 4264
Conc: 0.12 ng/ml



#9 O-Xylene

R.T.: 13.916 min
Delta R.T.: -0.009 min
Response: 5124
Conc: 0.15 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.208 min
Delta R.T.: 0.005 min
Response: 10894
Conc: 0.39 ng/ml