

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB011520\
 Data File : FB023414.D
 Signal(s) : FID2B.CH
 Acq On : 15 Jan 2020 9:14
 Operator : DD/AJ
 Sample : BSF0115S1
 Misc : 5.01g/5.00ml DI WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :
 BSF0115S1

Integration File: Calibration.e
 Quant Time: Jan 16 00:09:27 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB121919.M
 Quant Title :
 QLast Update : Thu Dec 19 12:43:06 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.794	494391	20.848 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.687	887511	28.676 ng/ml
2) t 2,2,4-Trimethylpentane	7.413	1031050	30.864 ng/ml
3) t n-Heptane	7.753	330739	10.027 ng/ml
4) t Benzene	7.887	405312	10.150 ng/ml
6) t Toluene	10.625	1172780	30.459 ng/ml
7) t Ethylbenzene	13.065	369631	10.184 ng/ml
8) t m-Xylene	13.198	702801	20.247 ng/ml
9) t O-Xylene	13.926	677198	20.105 ng/ml
10) t 1,2,4-Trimethylbenzene	16.205	579109	20.640 ng/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB011520\
Data File : FB023414.D
Signal(s) : FID2B.CH
Acq On : 15 Jan 2020 9:14
Operator : DD/AJ
Sample : BSF0115S1
Misc : 5.01g/5.00ml DI WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :
BSF0115S1

Integration File: Calibration.e
Quant Time: Jan 16 00:09:27 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB121919.M
Quant Title :
QLast Update : Thu Dec 19 12:43:06 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

