

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB011619\
 Data File : FB019426.D
 Signal(s) : FID2B.CH
 Acq On : 16 Jan 2019 12:27
 Operator : MJ/SJ
 Sample : K1010-03MSD
 Misc : 5.02g/5.00ml DI WATER
 ALS Vial : 10 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Jan 17 00:54:36 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB010819.M
 Quant Title :
 QLast Update : Wed Jan 09 07:19:31 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	9.042	176617	19.876 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.897	483224	24.263 ng/ml
2) t 2,2,4-Trimethylpentane	7.656	428022	22.468 ng/ml
3) t n-Heptane	7.998	132277	7.581 ng/ml
4) t Benzene	8.133	196597	10.012 ng/ml
6) t Toluene	10.890	546207	30.196 ng/ml
7) t Ethylbenzene	13.341	210062	9.481 ng/ml
8) t m-Xylene	13.474	439041	19.132 ng/ml
9) t O-Xylene	14.206	425398	18.959 ng/ml
10) t 1,2,4-Trimethylbenzene	16.493	397800	17.446 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB011619\
Data File : FB019426.D
Signal(s) : FID2B.CH
Acq On : 16 Jan 2019 12:27
Operator : MJ/SJ
Sample : K1010-03MSD
Misc : 5.02g/5.00ml DI WATER
ALS Vial : 10 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Jan 17 00:54:36 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB010819.M
Quant Title :
QLast Update : Wed Jan 09 07:19:31 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.897 min
Delta R.T.: 0.000 min
Response: 483224
Conc: 24.26 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.656 min
Delta R.T.: -0.001 min
Response: 428022
Conc: 22.47 ng/ml




#3 n-Heptane

R.T.: 7.998 min
Delta R.T.: 0.000 min
Response: 132277
Conc: 7.58 ng/ml




#4 Benzene


R.T.: 8.133 min
Delta R.T.: -0.002 min
Response: 196597
Conc: 10.01 ng/ml




#5 AAA-TFT
 R.T.: 9.042 min
 Delta R.T.: -0.002 min
 Response: 176617
 Conc: 19.88 ng/ml




#6 Toluene
 R.T.: 10.890 min
 Delta R.T.: 0.000 min
 Response: 546207
 Conc: 30.20 ng/ml



#7 Ethylbenzene
 R.T.: 13.341 min
 Delta R.T.: -0.001 min
 Response: 210062
 Conc: 9.48 ng/ml



#8 m-Xylene
 R.T.: 13.474 min
 Delta R.T.: -0.001 min
 Response: 439041
 Conc: 19.13 ng/ml




#9 O-Xylene

R.T.: 14.206 min

Delta R.T.: -0.001 min

Response: 425398

Conc: 18.96 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.493 min

Delta R.T.: -0.001 min

Response: 397800

Conc: 17.45 ng/ml