

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB011719\
 Data File : FB019453.D
 Signal(s) : FID2B.CH
 Acq On : 17 Jan 2019 12:12
 Operator : MJ/SJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Jan 18 02:47:57 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB010819.M
 Quant Title :
 QLast Update : Wed Jan 09 07:19:31 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	9.041	159838	17.988 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.894	507552	25.484 ng/ml
2) t 2,2,4-Trimethylpentane	7.656	523168	27.463 ng/ml
3) t n-Heptane	7.995	158394	9.078 ng/ml
4) t Benzene	8.133	189469	9.649 ng/ml
6) t Toluene	10.888	546747	30.225 ng/ml
7) t Ethylbenzene	13.339	222570	10.046 ng/ml
8) t m-Xylene	13.472	461944	20.130 ng/ml
9) t O-Xylene	14.204	451527	20.124 ng/ml
10) t 1,2,4-Trimethylbenzene	16.491	455060	19.957 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB011719\
Data File : FB019453.D
Signal(s) : FID2B.CH
Acq On : 17 Jan 2019 12:12
Operator : MJ/SJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 7 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Jan 18 02:47:57 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB010819.M
Quant Title :
QLast Update : Wed Jan 09 07:19:31 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.894 min
Delta R.T.: -0.003 min
Response: 507552
Conc: 25.48 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.656 min
Delta R.T.: -0.002 min
Response: 523168
Conc: 27.46 ng/ml




#3 n-Heptane

R.T.: 7.995 min
Delta R.T.: -0.003 min
Response: 158394
Conc: 9.08 ng/ml




#4 Benzene

R.T.: 8.133 min
Delta R.T.: -0.002 min
Response: 189469
Conc: 9.65 ng/ml




#5 AAA-TFT

R.T.: 9.041 min
Delta R.T.: -0.003 min
Response: 159838
Conc: 17.99 ng/ml




#6 Toluene

R.T.: 10.888 min
Delta R.T.: -0.003 min
Response: 546747
Conc: 30.23 ng/ml




#7 Ethylbenzene

R.T.: 13.339 min
Delta R.T.: -0.003 min
Response: 222570
Conc: 10.05 ng/ml



#8 m-Xylene

R.T.: 13.472 min
Delta R.T.: -0.004 min
Response: 461944
Conc: 20.13 ng/ml




#9 O-Xylene

R.T.: 14.204 min

Delta R.T.: -0.003 min

Response: 451527

Conc: 20.12 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.491 min

Delta R.T.: -0.003 min

Response: 455060

Conc: 19.96 ng/ml