

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB011823\
Data File : FB029124.D
Signal(s) : FID2B.CH
Acq On : 18 Jan 2023 10:46
Operator : AJ/MA
Sample : BSF0118S1
Misc : 5.00G/5.00 ML DI WATER
ALS Vial : 4 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Jan 18 23:23:40 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB011723.M
Quant Title :
QLast Update : Tue Jan 17 12:59:31 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

Compound		R.T.	Response	Conc Units

System Monitoring Compounds				
5) s	AAA-TFT	8.790	281269	23.018 ng/ml
Target Compounds				
1) t	2-Methylpentane	4.708	488221	33.347 ng/ml
2) t	2,2,4-Trimethylpentane	7.415	561937	35.417 ng/ml
3) t	n-Heptane	7.749	179973	10.771 ng/ml
4) t	Benzene	7.886	215261	10.125 ng/ml
6) t	Toluene	10.619	625050	31.011 ng/ml
7) t	Ethylbenzene	13.055	198895	10.370 ng/ml
8) t	m-Xylene	13.188	395574	21.143 ng/ml
9) t	O-Xylene	13.915	386782	21.494 ng/ml
10) t	1,2,4-Trimethylbenzene	16.190	394086	22.105 ng/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB011823\
Data File : FB029124.D
Signal(s) : FID2B.CH
Acq On : 18 Jan 2023 10:46
Operator : AJ/MA
Sample : BSF0118S1
Misc : 5.00G/5.00 ML DI WATER
ALS Vial : 4 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Jan 18 23:23:40 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB011723.M
Quant Title :
QLast Update : Tue Jan 17 12:59:31 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

