

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB011823\
 Data File : FB029128.D
 Signal(s) : FID2B.CH
 Acq On : 18 Jan 2023 14:09
 Operator : AJ/MA
 Sample : 01178-01MS 50X
 Misc : 2.84G/5.00 ML MEOH
 ALS Vial : 8 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Jan 18 23:23:59 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB011723.M
 Quant Title :
 QLast Update : Tue Jan 17 12:59:31 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.791	271837	22.246 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.705	398708	27.233 ng/ml
2) t 2,2,4-Trimethylpentane	7.420	482443	30.407 ng/ml
3) t n-Heptane	7.748	159392	9.539 ng/ml
4) t Benzene	7.885	211353	9.941 ng/ml
6) t Toluene	10.619	605667	30.050 ng/ml
7) t Ethylbenzene	13.056	189145	9.861 ng/ml
8) t m-Xylene	13.189	373815	19.980 ng/ml
9) t O-Xylene	13.916	362801	20.161 ng/ml
10) t 1,2,4-Trimethylbenzene	16.192	340308	19.089 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB011823\
Data File : FB029128.D
Signal(s) : FID2B.CH
Acq On : 18 Jan 2023 14:09
Operator : AJ/MA
Sample : 01178-01MS 50X
Misc : 2.84G/5.00 ML MEOH
ALS Vial : 8 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Jan 18 23:23:59 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB011723.M
Quant Title :
QLast Update : Tue Jan 17 12:59:31 2023
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.705 min
Delta R.T.: -0.002 min
Response: 398708
Conc: 27.23 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.420 min
Delta R.T.: 0.002 min
Response: 482443
Conc: 30.41 ng/ml




#3 n-Heptane

R.T.: 7.748 min
Delta R.T.: 0.000 min
Response: 159392
Conc: 9.54 ng/ml




#4 Benzene

R.T.: 7.885 min
Delta R.T.: 0.000 min
Response: 211353
Conc: 9.94 ng/ml




#5 AAA-TFT

R.T.: 8.791 min
Delta R.T.: 0.001 min
Response: 271837
Conc: 22.25 ng/ml




#6 Toluene

R.T.: 10.619 min
Delta R.T.: 0.001 min
Response: 605667
Conc: 30.05 ng/ml




#7 Ethylbenzene

R.T.: 13.056 min
Delta R.T.: 0.002 min
Response: 189145
Conc: 9.86 ng/ml



#8 m-Xylene

R.T.: 13.189 min
Delta R.T.: 0.002 min
Response: 373815
Conc: 19.98 ng/ml




#9 O-Xylene

R.T.: 13.916 min

Delta R.T.: 0.002 min

Response: 362801

Conc: 20.16 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.192 min

Delta R.T.: 0.002 min

Response: 340308

Conc: 19.09 ng/ml