

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB012119\
Data File : FB019481.D
Signal(s) : FID2B.CH
Acq On : 21 Jan 2019 14:12
Operator : MJ/SJ
Sample : K1009-05
Misc : 5.05g/5.00ml DI WATER
ALS Vial : 10 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Jan 22 06:34:14 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB010819.M
Quant Title :
QLast Update : Wed Jan 09 07:19:31 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	9.046	165988	18.680 ng/ml
Target Compounds			
8) t m-Xylene	13.477	7362	0.321 ng/ml
9) t O-Xylene	14.210	6740	0.300 ng/ml
10) t 1,2,4-Trimethylbenzene	16.508	5291	0.232 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB012119\
Data File : FB019481.D
Signal(s) : FID2B.CH
Acq On : 21 Jan 2019 14:12
Operator : MJ/SJ
Sample : K1009-05
Misc : 5.05g/5.00ml DI WATER
ALS Vial : 10 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Jan 22 06:34:14 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB010819.M
Quant Title :
QLast Update : Wed Jan 09 07:19:31 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#5 AAA-TFT

R.T.: 9.046 min
Delta R.T.: 0.002 min
Response: 165988
Conc: 18.68 ng/ml




#8 m-Xylene

R.T.: 13.477 min
Delta R.T.: 0.000 min
Response: 7362
Conc: 0.32 ng/ml



#9 O-Xylene

R.T.: 14.210 min
Delta R.T.: 0.003 min
Response: 6740
Conc: 0.30 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.508 min
Delta R.T.: 0.014 min
Response: 5291
Conc: 0.23 ng/ml