

Data Path : P:\HPCHEM1\FID_B\Data\FB012218\
 Data File : FB013151.D
 Signal(s) : FID2B.CH
 Acq On : 22 Jan 2018 12:04
 Operator : MJ/SJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :

Integration File: Calibration.e
 Quant Time: Jan 22 15:08:40 2018
 Quant Method : P:\HPCHEM1\FID_B\Method\FB012218.M
 Quant Title :
 QLast Update : Mon Jan 22 11:28:15 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	9.134	155431	20.375 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.980	400727	31.208 ng/ml
2) t 2,2,4-Trimethylpentane	7.748	379949	30.979 ng/ml
3) t n-Heptane	8.089	119860	10.331 ng/ml
4) t Benzene	8.229	156717	10.406 ng/ml
6) t Toluene	10.983	416128	30.328 ng/ml
7) t Ethylbenzene	13.432	163290	9.709 ng/ml
8) t m-Xylene	13.564	352846	19.455 ng/ml
9) t o-Xylene	14.296	374324	19.632 ng/ml
10) t 1,2,4-Trimethylbenzene	16.578	401329	20.358 ng/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : P:\HPCHEM1\FID_B\Data\FB012218\
Data File : FB013151.D
Signal(s) : FID2B.CH
Acq On : 22 Jan 2018 12:04
Operator : MJ/SJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :

Integration File: Calibration.e
Quant Time: Jan 22 15:08:40 2018
Quant Method : P:\HPCHEM1\FID_B\Method\FB012218.M
Quant Title :
QLast Update : Mon Jan 22 11:28:15 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane


R.T.: 4.980 min
Delta R.T.: -0.003 min
Response: 400727
Conc: 31.21 ng/ml

Instrument :
FID_B
ClientSampleId :




#2 2,2,4-Trimethylpentane

R.T.: 7.748 min
Delta R.T.: -0.002 min
Response: 379949
Conc: 30.98 ng/ml




#3 n-Heptane

R.T.: 8.089 min
Delta R.T.: -0.002 min
Response: 119860
Conc: 10.33 ng/ml



#4 Benzene


R.T.: 8.229 min
Delta R.T.: 0.000 min
Response: 156717
Conc: 10.41 ng/ml



#5 AAA-TFT


R.T.: 9.134 min
Delta R.T.: 0.000 min
Response: 155431
Conc: 20.38 ng/ml

Instrument :
FID_B
ClientSampleId :




#6 Toluene

R.T.: 10.983 min
Delta R.T.: 0.000 min
Response: 416128
Conc: 30.33 ng/ml




#7 Ethylbenzene

R.T.: 13.432 min
Delta R.T.: 0.000 min
Response: 163290
Conc: 9.71 ng/ml



#8 m-Xylene


R.T.: 13.564 min
Delta R.T.: 0.000 min
Response: 352846
Conc: 19.45 ng/ml



#9 O-Xylene

R.T.: 14.296 min
Delta R.T.: 0.000 min
Response: 374324
Conc: 19.63 ng/ml

Instrument :
FID_B
ClientSampleId :



#10 1,2,4-Trimethylbenzene

R.T.: 16.578 min
Delta R.T.: 0.000 min
Response: 401329
Conc: 20.36 ng/ml