

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB012220\
 Data File : FB023478.D
 Signal(s) : FID2B.CH
 Acq On : 22 Jan 2020 12:55
 Operator : DD/AJ
 Sample : L1216-06MSD
 Misc : 5.05g/5.00ml DI WATER
 ALS Vial : 9 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Jan 23 00:18:25 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB011620.M
 Quant Title :
 QLast Update : Thu Jan 16 11:42:55 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.793	519493	21.870 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.687	863706	27.332 ng/ml
2) t 2,2,4-Trimethylpentane	7.411	895624	26.378 ng/ml
3) t n-Heptane	7.752	290672	8.680 ng/ml
4) t Benzene	7.886	408398	10.000 ng/ml
6) t Toluene	10.624	1138731	28.815 ng/ml
7) t Ethylbenzene	13.064	348378	9.328 ng/ml
8) t m-Xylene	13.197	663691	18.664 ng/ml
9) t O-Xylene	13.925	644782	18.660 ng/ml
10) t 1,2,4-Trimethylbenzene	16.203	522949	17.884 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB012220\
Data File : FB023478.D
Signal(s) : FID2B.CH
Acq On : 22 Jan 2020 12:55
Operator : DD/AJ
Sample : L1216-06MSD
Misc : 5.05g/5.00ml DI WATER
ALS Vial : 9 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Jan 23 00:18:25 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB011620.M
Quant Title :
QLast Update : Thu Jan 16 11:42:55 2020
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.687 min
Delta R.T.: 0.000 min
Response: 863706
Conc: 27.33 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.411 min
Delta R.T.: -0.005 min
Response: 895624
Conc: 26.38 ng/ml




#3 n-Heptane

R.T.: 7.752 min
Delta R.T.: -0.004 min
Response: 290672
Conc: 8.68 ng/ml




#4 Benzene

R.T.: 7.886 min
Delta R.T.: -0.003 min
Response: 408398
Conc: 10.00 ng/ml




#5 AAA-TFT

R.T.: 8.793 min
Delta R.T.: -0.004 min
Response: 519493
Conc: 21.87 ng/ml




#6 Toluene

R.T.: 10.624 min
Delta R.T.: -0.004 min
Response: 1138731
Conc: 28.82 ng/ml




#7 Ethylbenzene

R.T.: 13.064 min
Delta R.T.: -0.004 min
Response: 348378
Conc: 9.33 ng/ml




#8 m-Xylene

R.T.: 13.197 min
Delta R.T.: -0.003 min
Response: 663691
Conc: 18.66 ng/ml



#9 O-Xylene

R.T.: 13.925 min
Delta R.T.: -0.004 min
Response: 644782
Conc: 18.66 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.203 min
Delta R.T.: -0.003 min
Response: 522949
Conc: 17.88 ng/ml