

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB012320\
 Data File : FB023486.D
 Signal(s) : FID2B.CH
 Acq On : 23 Jan 2020 14:44
 Operator : DD/AJ
 Sample : VBF0123W1
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Jan 24 00:12:44 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB011620.M
 Quant Title :
 QLast Update : Thu Jan 16 11:42:55 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.795	485162	20.425 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.681	3330	0.105 ng/ml
2) t 2,2,4-Trimethylpentane	7.414	2085	0.061 ng/ml
3) t n-Heptane	7.753	1202	0.036 ng/ml
6) t Toluene	10.627	5389	0.136 ng/ml
8) t m-Xylene	13.198	2556	0.072 ng/ml
9) t O-Xylene	13.912	1519	0.044 ng/ml
10) t 1,2,4-Trimethylbenzene	16.213	2410	0.082 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB012320\
Data File : FB023486.D
Signal(s) : FID2B.CH
Acq On : 23 Jan 2020 14:44
Operator : DD/AJ
Sample : VBF0123W1
Misc :
ALS Vial : 7 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Jan 24 00:12:44 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB011620.M
Quant Title :
QLast Update : Thu Jan 16 11:42:55 2020
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.681 min
Delta R.T.: -0.006 min
Response: 3330
Conc: 0.11 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.414 min
Delta R.T.: 0.000 min
Response: 2085
Conc: 0.06 ng/ml




#3 n-Heptane

R.T.: 7.753 min
Delta R.T.: -0.003 min
Response: 1202
Conc: 0.04 ng/ml




#5 AAA-TFT

R.T.: 8.795 min
Delta R.T.: -0.002 min
Response: 485162
Conc: 20.42 ng/ml




#6 Toluene

R.T.: 10.627 min
Delta R.T.: 0.000 min
Response: 5389
Conc: 0.14 ng/ml




#8 m-Xylene

R.T.: 13.198 min
Delta R.T.: -0.003 min
Response: 2556
Conc: 0.07 ng/ml



#9 O-Xylene

R.T.: 13.912 min
Delta R.T.: -0.016 min
Response: 1519
Conc: 0.04 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.213 min
Delta R.T.: 0.006 min
Response: 2410
Conc: 0.08 ng/ml