

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB012819\
 Data File : FB019562.D
 Signal(s) : FID2B.CH
 Acq On : 28 Jan 2019 13:47
 Operator : MJ/SJ
 Sample : K1255-01MSD
 Misc : 5.05g/5.00ml DI WATER
 ALS Vial : 11 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Jan 29 00:15:53 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB010819.M
 Quant Title :
 QLast Update : Wed Jan 09 07:19:31 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	9.047	174403	19.627 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.901	497788	24.994 ng/ml
2) t 2,2,4-Trimethylpentane	7.662	506791	26.603 ng/ml
3) t n-Heptane	8.002	152079	8.716 ng/ml
4) t Benzene	8.139	188507	9.600 ng/ml
6) t Toluene	10.894	524536	28.998 ng/ml
7) t Ethylbenzene	13.346	218679	9.870 ng/ml
8) t m-Xylene	13.479	455352	19.842 ng/ml
9) t O-Xylene	14.211	453255	20.201 ng/ml
10) t 1,2,4-Trimethylbenzene	16.497	455893	19.993 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB012819\
Data File : FB019562.D
Signal(s) : FID2B.CH
Acq On : 28 Jan 2019 13:47
Operator : MJ/SJ
Sample : K1255-01MSD
Misc : 5.05g/5.00ml DI WATER
ALS Vial : 11 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Jan 29 00:15:53 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB010819.M
Quant Title :
QLast Update : Wed Jan 09 07:19:31 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.901 min
Delta R.T.: 0.004 min
Response: 497788
Conc: 24.99 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.662 min
Delta R.T.: 0.004 min
Response: 506791
Conc: 26.60 ng/ml




#3 n-Heptane

R.T.: 8.002 min
Delta R.T.: 0.004 min
Response: 152079
Conc: 8.72 ng/ml




#4 Benzene

R.T.: 8.139 min
Delta R.T.: 0.004 min
Response: 188507
Conc: 9.60 ng/ml




#5 AAA-TFT

R.T.: 9.047 min
Delta R.T.: 0.003 min
Response: 174403
Conc: 19.63 ng/ml




#6 Toluene

R.T.: 10.894 min
Delta R.T.: 0.003 min
Response: 524536
Conc: 29.00 ng/ml




#7 Ethylbenzene

R.T.: 13.346 min
Delta R.T.: 0.004 min
Response: 218679
Conc: 9.87 ng/ml



#8 m-Xylene

R.T.: 13.479 min
Delta R.T.: 0.003 min
Response: 455352
Conc: 19.84 ng/ml




#9 O-Xylene

R.T.: 14.211 min

Delta R.T.: 0.003 min

Response: 453255

Conc: 20.20 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.497 min

Delta R.T.: 0.003 min

Response: 455893

Conc: 19.99 ng/ml