

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB012920\
 Data File : FB023575.D
 Signal(s) : FID2B.CH
 Acq On : 29 Jan 2020 19:45
 Operator : DD/AJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Jan 30 00:44:49 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB011620.M
 Quant Title :
 QLast Update : Thu Jan 16 11:42:55 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.797	490131	20.634 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.686	919601	29.100 ng/ml
2) t 2,2,4-Trimethylpentane	7.415	1036141	30.516 ng/ml
3) t n-Heptane	7.755	335011	10.004 ng/ml
4) t Benzene	7.889	405635	9.933 ng/ml
6) t Toluene	10.628	1194218	30.219 ng/ml
7) t Ethylbenzene	13.067	375117	10.044 ng/ml
8) t m-Xylene	13.201	721517	20.290 ng/ml
9) t O-Xylene	13.928	700318	20.268 ng/ml
10) t 1,2,4-Trimethylbenzene	16.206	589545	20.162 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB012920\
Data File : FB023575.D
Signal(s) : FID2B.CH
Acq On : 29 Jan 2020 19:45
Operator : DD/AJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 24 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Jan 30 00:44:49 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB011620.M
Quant Title :
QLast Update : Thu Jan 16 11:42:55 2020
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.686 min
Delta R.T.: 0.000 min
Response: 919601
Conc: 29.10 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.415 min
Delta R.T.: 0.000 min
Response: 1036141
Conc: 30.52 ng/ml




#3 n-Heptane

R.T.: 7.755 min
Delta R.T.: 0.000 min
Response: 335011
Conc: 10.00 ng/ml




#4 Benzene

R.T.: 7.889 min
Delta R.T.: 0.000 min
Response: 405635
Conc: 9.93 ng/ml




#5 AAA-TFT

R.T.: 8.797 min
Delta R.T.: 0.000 min
Response: 490131
Conc: 20.63 ng/ml




#6 Toluene

R.T.: 10.628 min
Delta R.T.: 0.000 min
Response: 1194218
Conc: 30.22 ng/ml




#7 Ethylbenzene

R.T.: 13.067 min
Delta R.T.: 0.000 min
Response: 375117
Conc: 10.04 ng/ml




#8 m-Xylene

R.T.: 13.201 min
Delta R.T.: 0.000 min
Response: 721517
Conc: 20.29 ng/ml



#9 O-Xylene

R.T.: 13.928 min
Delta R.T.: 0.000 min
Response: 700318
Conc: 20.27 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.206 min
Delta R.T.: 0.000 min
Response: 589545
Conc: 20.16 ng/ml