

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB013019\
 Data File : FB019580.D
 Signal(s) : FID2B.CH
 Acq On : 30 Jan 2019 8:26
 Operator : MJ/SJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Jan 30 23:43:14 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB012919.M
 Quant Title :
 QLast Update : Tue Jan 29 12:07:12 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	9.050	202667	17.337 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.903	505097	29.174 ng/ml
2) t 2,2,4-Trimethylpentane	7.664	491870	29.660 ng/ml
3) t n-Heptane	8.004	157660	9.784 ng/ml
4) t Benzene	8.142	181816	9.639 ng/ml
6) t Toluene	10.897	494028	28.242 ng/ml
7) t Ethylbenzene	13.349	206146	9.393 ng/ml
8) t m-Xylene	13.482	430835	18.824 ng/ml
9) t O-Xylene	14.214	430331	19.267 ng/ml
10) t 1,2,4-Trimethylbenzene	16.501	448145	19.342 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB013019\
Data File : FB019580.D
Signal(s) : FID2B.CH
Acq On : 30 Jan 2019 8:26
Operator : MJ/SJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 1 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Jan 30 23:43:14 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB012919.M
Quant Title :
QLast Update : Tue Jan 29 12:07:12 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.903 min
Delta R.T.: 0.001 min
Response: 505097
Conc: 29.17 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.664 min
Delta R.T.: 0.000 min
Response: 491870
Conc: 29.66 ng/ml




#3 n-Heptane

R.T.: 8.004 min
Delta R.T.: 0.000 min
Response: 157660
Conc: 9.78 ng/ml




#4 Benzene

R.T.: 8.142 min
Delta R.T.: 0.000 min
Response: 181816
Conc: 9.64 ng/ml




#5 AAA-TFT

R.T.: 9.050 min
Delta R.T.: 0.000 min
Response: 202667
Conc: 17.34 ng/ml




#6 Toluene

R.T.: 10.897 min
Delta R.T.: 0.000 min
Response: 494028
Conc: 28.24 ng/ml




#7 Ethylbenzene

R.T.: 13.349 min
Delta R.T.: 0.000 min
Response: 206146
Conc: 9.39 ng/ml



#8 m-Xylene

R.T.: 13.482 min
Delta R.T.: 0.000 min
Response: 430835
Conc: 18.82 ng/ml




#9 O-Xylene

R.T.: 14.214 min

Delta R.T.: 0.000 min

Response: 430331

Conc: 19.27 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.501 min

Delta R.T.: 0.000 min

Response: 448145

Conc: 19.34 ng/ml