

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB013120\
 Data File : FB023592.D
 Signal(s) : FID2B.CH
 Acq On : 31 Jan 2020 8:41
 Operator : DD/AJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Jan 31 22:35:47 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB011620.M
 Quant Title :
 QLast Update : Thu Jan 16 11:42:55 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.788	485548	20.441 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.682	846997	26.803 ng/ml
2) t 2,2,4-Trimethylpentane	7.407	917910	27.034 ng/ml
3) t n-Heptane	7.747	301838	9.014 ng/ml
4) t Benzene	7.881	404241	9.899 ng/ml
6) t Toluene	10.620	1151068	29.127 ng/ml
7) t Ethylbenzene	13.061	356962	9.558 ng/ml
8) t m-Xylene	13.194	684473	19.248 ng/ml
9) t O-Xylene	13.922	662360	19.169 ng/ml
10) t 1,2,4-Trimethylbenzene	16.201	554856	18.975 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB013120\
Data File : FB023592.D
Signal(s) : FID2B.CH
Acq On : 31 Jan 2020 8:41
Operator : DD/AJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 1 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Jan 31 22:35:47 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB011620.M
Quant Title :
QLast Update : Thu Jan 16 11:42:55 2020
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.682 min
Delta R.T.: -0.005 min
Response: 846997
Conc: 26.80 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.407 min
Delta R.T.: -0.008 min
Response: 917910
Conc: 27.03 ng/ml




#3 n-Heptane

R.T.: 7.747 min
Delta R.T.: -0.009 min
Response: 301838
Conc: 9.01 ng/ml




#4 Benzene

R.T.: 7.881 min
Delta R.T.: -0.008 min
Response: 404241
Conc: 9.90 ng/ml




#5 AAA-TFT

R.T.: 8.788 min
 Delta R.T.: -0.008 min
 Response: 485548
 Conc: 20.44 ng/ml




#6 Toluene

R.T.: 10.620 min
 Delta R.T.: -0.008 min
 Response: 1151068
 Conc: 29.13 ng/ml




#7 Ethylbenzene

R.T.: 13.061 min
 Delta R.T.: -0.007 min
 Response: 356962
 Conc: 9.56 ng/ml




#8 m-Xylene

R.T.: 13.194 min
 Delta R.T.: -0.007 min
 Response: 684473
 Conc: 19.25 ng/ml



#9 O-Xylene

R.T.: 13.922 min
Delta R.T.: -0.007 min
Response: 662360
Conc: 19.17 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.201 min
Delta R.T.: -0.006 min
Response: 554856
Conc: 18.98 ng/ml