

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB020123\
Data File : FB029203.D
Signal(s) : FID2B.CH
Acq On : 1 Feb 2023 16:21
Operator : AJ/MA
Sample : BSF0201S2
Misc : 5.00G/5.00 ML DI WATER
ALS Vial : 9 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Feb 01 16:30:23 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB013123.M
Quant Title :
QLast Update : Tue Jan 31 12:31:55 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 t	2-Methylpentane	11.492	10.993 E3	4.3	102	0.00
2 t	2,2,4-Trimethylpentane	14.158	13.485 E3	4.8	99	0.00
3 t	n-Heptane	12.441	11.998 E3	3.6	105	0.00
4 t	Benzene	18.021	18.122 E3	-0.6	112	0.00
5 s	AAA-TFT	11.937	9.996 E3	16.3	108	0.00
6 t	Toluene	17.535	18.282 E3	-4.3	116	0.00
7 t	Ethylbenzene	16.900	17.411 E3	-3.0	113	0.00
8 t	m-Xylene	16.754	17.220 E3	-2.8	112	0.00
9 t	O-Xylene	16.608	16.737 E3	-0.8	109	0.00
10 t	1,2,4-Trimethylbenzene	16.961	16.759 E3	1.2	105	0.00

Evaluate Continuing Calibration Report - Not Found

(#) = Out of Range

SPCC's out = 0 CCC's out = 0